

**Amendment No.1**

Date: 20.07.2018

**Subject: Amendment no. 1 to the Tender Enquiry Document****Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/27-RT/18-19 Dated: 11.07.2018**

The pre bid meeting of the above referred tender enquiry was held on 17.07.2018. Based on pre-bid discussions following amendments are being incorporated in the tender enquiry document.

**SECTION I**  
**NOTICE INVITING TENDER (NIT)**

**For:-**

Sl No	RFx No	Equipment	Department	Quantity per AIIMS	Qty. for Six AIIMS	Tender Processing Fee (INR)*	EMD (INR)
1	3000002961	External Defibrillator	Cardiology	5	30	3,540	3,00,000

**Read As:-**

Sl No	RFx No	Equipment	Department	Quantity per AIIMS	Qty. for Six AIIMS	Tender Processing Fee (INR)*	EMD (INR)
1	<b>3000002965</b>	External Defibrillator	Cardiology	5	30	3,540	3,00,000

**Section VII**  
**Technical Specifications**

<b>Sch 01. External Defibrillator (Rfx no. 3000002965)</b>		
<b>Tender Page No. &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
<b>Page 45 Para 7</b>	The device should be AC/12V DC powered with ability to run on internal battery and power battery should provide a minimum of 3 hours continuous monitoring and at least 25 full voltage shocks. Defibrillators working only on battery should not be quoted.	The device should be AC/12V DC powered with ability to run on internal battery and power battery should provide a minimum of 3 hours continuous monitoring and <b>at least 50</b> full voltage shocks. Defibrillators working only on battery should not be quoted.
<b>Page 45 Para 8</b>	Should have ability to display record and print at least 3- channel ECG simultaneously.	Should have ability to display record and print <b>at least 2- channel</b> ECG simultaneously.

<b>Page 45 Para 10</b>	Should have connection for attaching disposable patch electrodes from which pacing /defibrillation /ECG recording is possible. Vendor has to provide connecting cables for at least 3 commonly available patches available in the market.	Should have connection for attaching disposable patch electrodes from which pacing /defibrillation /ECG recording is possible. Vendor has to provide connecting cables for <b>at least 3 OEM patches.</b>
<b>Page 45 Para 14</b>	Vendor has to provide the following accessories with each unit – no's are given in brackets : reusable NIBP cuffs ( all 4 sizes , from neonate to adult ,5 each ) Disposable NIBP cuffs for neonates (25),reusable SpO2 probes(adult ,pediatric ,infant and neonate-5 each), disposable SpO2 probes for neonates and infants (50), Disposable adult and pediatric defibrillation patches (10 each) as already mentioned in 10. Prices of all the accessories should be quoted separately and should be frozen for next 5 years	Vendor has to provide the following accessories with each unit – no's are given in brackets: reusable NIBP cuffs (all 4 sizes, from neonate to adult, <b>3 each</b> ) Disposable NIBP cuffs for neonates (25), reusable SpO2 probes (adult, pediatric, infant <b>or</b> neonate - <b>3 each</b> ), disposable SpO2 probes for neonates and infants (50), Disposable adult and pediatric defibrillation patches ( <b>3 each</b> ) as already mentioned in 10. Prices of all the accessories should be quoted separately and should be frozen for next 5 years
<b>Page 46 Para 19</b>	<b>Optional</b> CPR measurements and feedback: Compression depth, rate, ventilation rate. Provide appropriate cables and sensors (reusable 10 each). Include cost of above in main bid but list this separately.	CPR measurements and feedback: Compression depth, rate, ventilation rate. Provide appropriate cables and sensors (reusable <b>05 each</b> ). Include cost of above in main bid but list this separately. ( <b>Standard items to be quoted mandatorily</b> )

**Sch 02. Urodynamic System (Rfx no. 3000002962)**

<b>Tender Page No. &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
<b>Page 46 Para 3</b>	Equipment should have 5-8 channels	Equipment should have <b>at least 3 pressure</b> channels
<b>Page 47 Para 20</b>	2 Lumen Catheter 8 fr - 50 nos	2 Lumen Catheter <b>7/8 fr - 40 nos and 5/6 fr - 10 nos</b>
<b>Page 47 Para 26</b>	Should be compatible for Hospital Information system	<b>Deleted</b>
<b>Page 47 Para 29</b>	Should have automatic artifact recognition and compensation on flow transducer.	<b>Deleted</b>

**Sch 03. Holmium Laser 120W (Rfx no. 3000002963)**

<b>Tender Page No. &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
<b>Page 49 Para 20 viii</b>	Single Chamber Imported Tissue Evacuator – 1	<b>Deleted</b>

<b>Sch 04. Laparoscopy Unit (Rfx no. 3000002964)</b>		
<b>Tender Page No. &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
<b>Page 49 Para C 3</b>	<b>Light Source: (01)</b> "LED light source for general purpose use and ICG. If LED not available for ICG purpose the vendor may offer LED source for general purpose and xenon source compatible with ICG"	<b>LED/Xenon Light source for general purpose and ICG use. LED/Xenon bulb will be covered under warranty for 5 years.</b>
<b>Page 51 Para G 5</b>	<b>" Essential accessories (Of same manufacturer as main equipment and 15% of tolerance is acceptable in sizes mention below)</b> Right angled dissector with smooth curve: laparoscopic 10mm, 34-36cm, insulated metal handle and sheath, gun-style handle, rotatable, without ratchet, reusable (02)	Right angled dissector with smooth curve: laparoscopic <b>5/10mm</b> , 34-36cm, insulated metal handle and sheath, gun-style handle, rotatable, without ratchet, reusable (02)

**All other terms and conditions of the tender enquiry remain unaltered.**