

M/s. HLL Infra Tech Services Ltd. (HITES)
(Subsidiary of HLL Infra tech Services Ltd., A Government of India Enterprise)
For & on behalf of
U. N. Mehta Institute of cardiology & Research Centre,
(Affiliated to B. J. Medical College, Ahmedabad, Gujarat)

CORRIGENDUM – 01 (ADDENDUM)

Date: 02.03.2019

Name of the Tender: e-TENDER FOR PROCUREMENT OF SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF HIGHLY TECHNICAL AND HIGHLY SOPHISTICATED MEDICAL EQUIPMENT FOR VARIOUS DEPARTMENT FOR NEW CONSTRUCTION OF PHASE-II, PAEDIATRIC CARDIAC HOSPITAL BUILDING (GRIHA 3 STAR RATED) BY HLL INFRATECH SERVICES LTD. ON BEHALF OF U. N. MEHTA INSTITUTE OF CARDIOLOGY & RESEARCH CENTRE, AHMEDABAD, GUJARAT

Reference: e-TENDER ID # 347945

The Corrigendum – 01(Addendum) shall be treated as part of e-tender. Scanned copy of the same to be uploaded online along with the other e-tender documents and print out of uploaded documents to be submitted in the form of hard copy along with submission of e-tender, which should be duly signed and stamped:

Amendment to existing e-tender terms & conditions:

SN	Reference to E-Tender Clause	Amended Clause (New Clause Added)
1	Section-VII: Technical Specification: Item no.:4: Point-2.8: Lockable door with front glass and tight sealing (Magnetic closing) surround to prevent cold loss. Should have at least 4 rollout type drawers with stainless steel make	Section-VII: Technical Specification: Item no.:4: Point-2.8: Lockable door with front glass and tight sealing, surround to prevent cold loss. Should have at least 4 rollout type drawers with stainless steel make
2	Section-VII: Technical Specification: Item no.:4: Point-2.9: Automatic defrosting and condensed melt water evaporation.	Deleted

e-TENDER FOR PROCUREMENT OF SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF HIGHLY TECHNICAL AND HIGHLY SOPHISTICATED MEDICAL EQUIPMENT FOR VARIOUS DEPARTMENT

SN	Reference to E-Tender Clause	Amended Clause (New Clause Added)
3	Section-VII: Technical Specification: Item no.:4: Point-2.13: Epoxy coated outside finish and GS interior.	Section-VII: Technical Specification: Item no.:4: Point-2.13: Epoxy coated/ MS powder coated outside finish and SS 304 interior.
4	Section-VII: Technical Specification: Item no.:4: Point-2.14: Low noise, automatic defrosting, CFC free & HCFC free.	Section-VII: Technical Specification: Item no.:4: Point-2.14: Low noise, CFC free & HCFC free.
5	Section-VII: Technical Specification: Item no.:7: Point-5: Should deliver constant pre-set pressure ranges – Distal 40 - 160 mm Hg	Section-VII: Technical Specification: Item no.:7: Point-5: Should deliver constant pre-set pressure ranges – Distal 40 - 130 mm Hg (± 5 mm Hg)
6	Section-VII: Technical Specification: Item no.:5: Point-A (1.3): Equipped with special detection technology for detailed observation by enhancing visibility of blood capillaries & mucosa	Section-VII: Technical Specification: Item no.:5: Point-A (1.3): Equipped with Optical technology for detailed observation by enhancing visibility of blood capillaries & mucosa
7	Section-VII: Technical Specification: Item no.:5: Point-A (1.4): Equipped with internal & external portable memory & its slot for image recording.	Section-VII: Technical Specification: Item no.:5: Point-A (1.4): Equipped with internal or external portable memory & its slot for image recording.
8	Section-VII: Technical Specification: Item no.:5: Point-A (1.7): Electronic zoom upto 1.5x	Section-VII: Technical Specification: Item no.:5: Point-A (1.7): Electronic zoom upto 1.5x or better
9	Section-VII: Technical Specification: Item no.:5: Point-A (1.9): Backlit front panel indicators	Section-VII: Technical Specification: Item no.:5: Point-A (1.9): Backlit front panel indicators or LED Screen
10	Section-VII: Technical Specification: Item no.:5: Point-B (2.2): Backlit front panel indicators	Section-VII: Technical Specification: Item no.:5: Point-B (2.2): Backlit front panel indicators or LED Screen
11	Section-VII: Technical Specification: Item no.:5: Point-D (7) & D(8): Distal end outer diameter: 9.8 mm or less and Insertion tube outer diameter : 9.8 mm or less	Section-VII: Technical Specification: Item no.:5: Point-D (7) & D(8): Distal end outer diameter: 9.9 mm or less and Insertion tube outer diameter : 9.9 mm or less
12	Section-VII: Technical Specification: Item no.:5: Point-F: Technical Specification of Video Bronchoscope Adult	Section-VII: Technical Specification: Item no.:5: Point-F: Technical Specification of Video Bronchoscope Adult, It should be High Resolution or High Definition.

e-TENDER FOR PROCUREMENT OF SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF HIGHLY TECHNICAL AND HIGHLY SOPHISTICATED MEDICAL EQUIPMENT FOR VARIOUS DEPARTMENT

SN	Reference to E-Tender Clause	Amended Clause (New Clause Added)
13	Section-VII: Technical Specification: Item no.:5: Point-F (3): Distal end diameter should be 5.2 mm or less	Section-VII: Technical Specification: Item no.:5: Point-F (3): Distal end diameter should be 5.5 mm or less
14	Section-VII: Technical Specification: Item no.:5: Point-G: Technical Specification of Video Bronchoscope Paediatric	Section-VII: Technical Specification: Item no.:5: Point-G: Technical Specification of Video Bronchoscope Paediatric, It should be High Resolution or High Definition
15	Section-VII: Technical Specification: Item no.:5: Point-H (4): Field of view : 140 degree or more	Section-VII: Technical Specification: Item no.:5: Point-H (4): Field of view : 120 degree or more
16	Section-VII: Technical Specification: Item no.:5: Point-H (6): Depth of field : 3 to 100 mm or better	Section-VII: Technical Specification: Item no.:5: Point-H (6): Depth of field : 4 to 100 mm or better
17	Section-VII: Technical Specification: Item no.:5: Point-D (2): Should be suitable for special detection technology for detailed observation by enhancing visibility of blood capillaries & mucosa	Section-VII: Technical Specification: Item no.:5: Point-D (2): Should be suitable for Optical technology for detailed observation by enhancing visibility of blood capillaries & mucosa
18	Section-VII: Technical Specification: Item no.:5: Point-E (2): Suitable for special detection technology for detailed observation by enhancing visibility of blood capillaries & mucosa	Section-VII: Technical Specification: Item no.:5: Point-E (2): Should be suitable for Optical technology for detailed observation by enhancing visibility of blood capillaries & mucosa
19	Section-VII: Technical Specification: Item no.:5: Point-H: Suitable for special detection technology for detailed observation by enhancing visibility of blood capillaries & mucosa	Section-VII: Technical Specification: Item no.:5: Point-H: Should be Suitable for Optical technology for detailed observation by enhancing visibility of blood capillaries & mucosa
20	Section-VII: Technical Specification: Item no.:10: Point-4: Should be supplied with heated breathing tube, auto fill chamber and adaptors (10 -pack)	Section-VII: Technical Specification: Item no.:10: Point-4: Should be supplied with heated breathing tube, auto fill chamber and adapters/ Connector (10 -pack)
21	Section-VII: Technical Specification: Item no.:10: Point-5: Should have audio visual alarm for leak, blockage, low O ₂ , High O ₂ and low temperature	Section-VII: Technical Specification: Item no.:10: Point-5: Should have audio visual alarm for leak, blockage, low O ₂ , High O ₂ and hardware fault with error codes.
22	Section-VII: Technical Specification: Item no.:10: Point-6: Serial port/USB port available for interface	Deleted

e-TENDER FOR PROCUREMENT OF SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF HIGHLY TECHNICAL AND HIGHLY SOPHISTICATED MEDICAL EQUIPMENT FOR VARIOUS DEPARTMENT

SN	Reference to E-Tender Clause	Amended Clause (New Clause Added)																														
23	Section-VII: Technical Specification: Item no.:10: Point-9(g): Mask Interface Adapter -22 mm (20 -pack)	Deleted																														
20	<p>SECTION I</p> <p>NOTICE INVITING TENDERS (NIT)</p> <p>(A) E tender Enquiry No.: HITES/PCD/UNM-AHD/Equ-Mix/18-19 Dated: 06.02.2019</p> <p>Important Dates and Venue for Events:</p> <table border="1" data-bbox="275 647 1137 1394"> <thead> <tr> <th data-bbox="275 647 360 719">SI. No.</th> <th data-bbox="360 647 741 719">Description</th> <th data-bbox="741 647 1137 719">Schedule</th> </tr> </thead> <tbody> <tr> <td data-bbox="275 719 360 799">i.</td> <td data-bbox="360 719 741 799">Pre-bid Meeting Date & Time</td> <td data-bbox="741 719 1137 799">13.02.2019, 10.00 A.M (Morning)</td> </tr> <tr> <td data-bbox="275 799 360 1098">ii.</td> <td data-bbox="360 799 741 1098">Pre-bid Meeting Venue</td> <td data-bbox="741 799 1137 1098">4th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.</td> </tr> <tr> <td data-bbox="275 1098 360 1209">iii.</td> <td data-bbox="360 1098 741 1209">Due date & time for submission of e tender in e tender Portal</td> <td data-bbox="741 1098 1137 1209">06.03.2019, 12.00 P.M (Noon)</td> </tr> <tr> <td data-bbox="275 1209 360 1394">iv.</td> <td data-bbox="360 1209 741 1394">Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form</td> <td data-bbox="741 1209 1137 1394">06.03.2019, 2.00 P.M (After Noon)</td> </tr> </tbody> </table>	SI. No.	Description	Schedule	i.	Pre-bid Meeting Date & Time	13.02.2019, 10.00 A.M (Morning)	ii.	Pre-bid Meeting Venue	4 th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.	iii.	Due date & time for submission of e tender in e tender Portal	06.03.2019, 12.00 P.M (Noon)	iv.	Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form	06.03.2019, 2.00 P.M (After Noon)	<p>SECTION I</p> <p>NOTICE INVITING TENDERS (NIT)</p> <p>(A) E tender Enquiry No.: HITES/PCD/UNM-AHD/Equ-Mix/18-19 Dated: 06.02.2019</p> <p>Important Dates and Venue for Events:</p> <table border="1" data-bbox="1160 647 2067 1394"> <thead> <tr> <th data-bbox="1160 647 1245 719">SI. No.</th> <th data-bbox="1245 647 1637 719">Description</th> <th data-bbox="1637 647 2067 719">Schedule</th> </tr> </thead> <tbody> <tr> <td data-bbox="1160 719 1245 799">i.</td> <td data-bbox="1245 719 1637 799">Pre-bid Meeting Date & Time</td> <td data-bbox="1637 719 2067 799">13.02.2019, 10.00 A.M (Morning)</td> </tr> <tr> <td data-bbox="1160 799 1245 1098">ii.</td> <td data-bbox="1245 799 1637 1098">Pre-bid Meeting Venue</td> <td data-bbox="1637 799 2067 1098">4th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.</td> </tr> <tr> <td data-bbox="1160 1098 1245 1209">iii.</td> <td data-bbox="1245 1098 1637 1209">Due date & time for submission of e tender in e tender Portal</td> <td data-bbox="1637 1098 2067 1209">19.03.2019, 12.00 P.M (Noon)</td> </tr> <tr> <td data-bbox="1160 1209 1245 1394">iv.</td> <td data-bbox="1245 1209 1637 1394">Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form</td> <td data-bbox="1637 1209 2067 1394">19.03.2019, 2.00 P.M (After Noon)</td> </tr> </tbody> </table>	SI. No.	Description	Schedule	i.	Pre-bid Meeting Date & Time	13.02.2019, 10.00 A.M (Morning)	ii.	Pre-bid Meeting Venue	4 th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.	iii.	Due date & time for submission of e tender in e tender Portal	19.03.2019, 12.00 P.M (Noon)	iv.	Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form	19.03.2019, 2.00 P.M (After Noon)
SI. No.	Description	Schedule																														
i.	Pre-bid Meeting Date & Time	13.02.2019, 10.00 A.M (Morning)																														
ii.	Pre-bid Meeting Venue	4 th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.																														
iii.	Due date & time for submission of e tender in e tender Portal	06.03.2019, 12.00 P.M (Noon)																														
iv.	Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form	06.03.2019, 2.00 P.M (After Noon)																														
SI. No.	Description	Schedule																														
i.	Pre-bid Meeting Date & Time	13.02.2019, 10.00 A.M (Morning)																														
ii.	Pre-bid Meeting Venue	4 th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.																														
iii.	Due date & time for submission of e tender in e tender Portal	19.03.2019, 12.00 P.M (Noon)																														
iv.	Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form	19.03.2019, 2.00 P.M (After Noon)																														

e-TENDER FOR PROCUREMENT OF SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF HIGHLY TECHNICAL AND HIGHLY SOPHISTICATED MEDICAL EQUIPMENT FOR VARIOUS DEPARTMENT

SN	Reference to E-Tender Clause			Amended Clause (New Clause Added)		
	v.	Date and Time of opening of Techno - Commercial e tenders	06.03.2019, 2.30 P.M (After noon)	v.	Date and Time of opening of Techno - Commercial e tenders	19.03.2019, 2.30 P.M (After noon)
	vi.	Venue for opening of E tenders	M/s. HLL Infra Tech Services Ltd., B14-A, Sector-62, Noida (UP) – 201307 India.	vi.	Venue for opening of E tenders	M/s. HLL Infra Tech Services Ltd., B14-A, Sector-62, Noida (UP) – 201307 India.

***** End of Corrigendum – 01 (Addendum) *****