

09-09-2020

Amendment No. 11**Sub: Amendment to the referred tender enquiry**

Ref.: HITES/PCD/AIIMS-IV/27/ENT/19-20 dated 29-01-2020 read with clarification no 1, 2, amendment no 2, 3, 4, 5, 6, 7,8 & Notice on Extension of Due Dates dated 04.03.2020, 13.03.2020, 03.03.2020, 18.03.2020, , 24.06.2020, 09.07.2020 & 16.07.2020, 30.07.2020, 10.08.2020, 17.08.2020 & 25.03.2020 respectively

The following changes are being incorporated in the following Tender ID only:

2020_HLL_41506_18

Tender Timeline for tender ID 2020_HLL_41506_18 (Endoscopic sinus surgery set Xenon head light with camera system)

SECTION I

NOTICE INVITING e-TENDER (e-NIT)

Existing:

Item No.	Tender ID	Name of the Item	Qty.	Unit	Tender processing Fee (Incl of GST)	EMD Amount
6	2020_HLL_41506_18	Endoscopic sinus surgery set Xenon head light with camera system	11	Set	₹ 2,950	₹ 22,00,000

Read as:

Item No.	Tender ID	Name of the Item	Qty.	Unit	Tender processing Fee (Incl of GST)	EMD Amount
6	2020_HLL_41506_18	Endoscopic sinus surgery set	11	Set	₹ 2,950	₹ 22,00,000

2) Tender Timeline:**Existing:**

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	10-09-2020, 1200 hrs
d.	Closing date & time for submission of tender processing fee and EMD in physical form*	11-09-2020, 1400 hrs
e.	Time and date of opening of online bids	11-09-2020, 1430 hrs

Read as:

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	16-09-2020, 1200 hrs
d.	Closing date & time for submission of tender processing fee and EMD in physical form*	17-09-2020, 1400 hrs
e.	Time and date of opening of online bids	17-09-2020, 1430 hrs

SECTION - VI
LIST OF REQUIREMENTS

Part I

Existing:

Item No.	Tender ID	Name of the Item	Qty.	Unit	Warranty Period	CMC Period
6	2020_HLL_41506_18	Endoscopic sinus surgery set Xenon head light with camera system	11	Set	5 years	5 years

Read as:

Item No.	Tender ID	Name of the Item	Qty.	Unit	Warranty Period	CMC Period
6	2020_HLL_41506_18	Endoscopic sinus surgery set	11	Set	5 years	5 years

Section VII
Technical Specification

Technical Specification of item no 6 (Endoscopic sinus surgery set) may be read as :

Sl. No	Technical Specification	Quantity
1	0 degree, 4 mm, 16-18cm wide angle straight forward telescope autoclavable	3
2	30 degree, 4 mm, 16-18cm wide angle straight forward telescope autoclavable	2
3	45 degree, 4 mm, 16-18cm wide angle straight forward telescope autoclavable	2
4	70 degree, 4 mm, 16-18cm wide angle straight forward telescope autoclavable	2
5	0 degree, 2.7 mm, 16-18cm straight forward telescope autoclavable	1
6	30 degree, 2.7 mm, 16-18cm straight forward telescope autoclavable	1
7	Deleted	2
8	Sickle knife, pointed, 19cm long	4
9	Freer elevator should be double ended, semi sharp and blunt, 20cm long	4
10	Small size oblong shaped Antrum curette, straight, 19cm	1
11	Sinus curette 90deg and 55 deg curved	02 each
12	Antrum curette forward cutting small size, 19cm length	1
13	Double ended maxillary sinus ostium seeker, ball shaped ends diameter 1.2 and 2mm, length 19cm	2
14	Cottle elevator double ended, semi sharp and blunt, graduated, length 20cm	4
15	Conical suction tube should be malleable, with finger grip plate, luer lock, Outer diameter 2.5cm, working length 13cm	1
16	Antrum cannula, Luer-lock, with cut-off hole, short curved, outer diameter 3mm & 4 mm, length 12.5cm	1
17	Bipolar coagulation forceps, insulated, angular, blunt, with integrated suction channel, with cut-off hole, length 19cm, to be supplied with Bipolar High frequency cord	1
18	Bipolar Suction forceps, 15 deg upturned, with suction channel, working length 12.5cm, to be supplied with Bipolar High frequency cord	2
19	Antrum punch for Left & Right side downward and forward cutting, working length 10cm	2
20	Nasal cutting forceps, working length 13cm	2
21	Antrum Punch, right and left side backward cutting, working length 10cm	2
22	Antrum grasping forceps for maxillary sinus, jaws curved to right, fixed jaw curved 90 deg, opening 120 deg, movable jaw backward, length 10cm	1
23	Blakesley nasal forceps, straight with working length 13cm	4
24	Blakesley nasal forceps, Upturned 45deg & 90deg with working length 13cm	02 each
25	Giraffe forceps 65deg upturn, cup jaws diameter 3mm with horizontal & vertical opening, length 12cm	2
26	Biopsy & Grasping forceps, vertical opening, malleable sheath end, cupped jaws diameter 4mm, working length 18cm	2
27	Sphenoid Punch, circular cutting circular punch, dia 4.5mm, working length 18cm	2
28	Sphenoid Punch, 65 deg upturned, circular cutting, dia 3.5mm, length 17cm	2
29	Nasal Scissors (Straight, right & left)	1

Sl. No	Technical Specification	Quantity
30	Antrum punch(small) Pediatric size , backward cutting	1
31	Biopsy forceps for nasopharynx	2
32	Turbinectomy Scissors	2
33	Tilley henckel forceps	4
34	Through cut forceps(straight/ 45deg) – 18x3mm, 11.5x3.5mm	02 each
35	Malleable suction	2
36	DCR punch	2
37	Frontal sinus seeker(Double ended 22cm- 70deg and 90deg)	1
	Deleted	
38	All components should be from same principal manufacturer. All the quoted components should be USFDA or European CE with 4 digit notified body number or BIS certified.	

All other contents of the Tender enquiry including terms & conditions remain unaltered.

Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids
- II. Technical Compliance Sheet ‘Ver 1.1’ and Price bid Sheet ‘Ver 1.0’ in excel form against referred tender ID 2020_HLL_41506_18 are being uploaded under ‘Corrigendum’ on CPPP. The same need to be downloaded for submission