

14-05-2018

Amendment No. 3**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/30/URLG/18-19 dated 10/04/2018
Amendment No. 01 dated 28-04-2018
Amendment No. 02 dated 08-05-2018**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**SECTION VII
TECHNICAL SPECIFICATION**

| Schedule No. 13 OT Table- Urology (Rfx number 3000002775) | | |
|---|-------------------------------------|------------------------------------|
| Para | CURRENT TENDER SPECIFICATION | COMMITTEE RECOMMENDATION |
| c.12 | Longitudinal Slide: atleast 300 mm | Longitudinal Slide: atleast 250 mm |

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**