

16-10-2018

Amendment No. 8**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/30/URLG/18-19 dated 10/04/2018****Amendment No. 01 dated 28-04-2018****Amendment No. 02 dated 08-05-2018****Amendment No. 03 dated 14-05-2018****Amendment No. 04 dated 08-06-2018****Amendment No. 05 dated 09-07-2018****Amendment No. 06 dated 09-08-2018****Amendment No. 07**

The following changes are being incorporated in the referred tender enquiry for the said item.

Section VII
Technical Specification

Schedule No. 02 ESWL (Rfx number 3000002764)		
Para	Current tender specification	Read as
	Suitable UPS should be supplied with 1hrs backup approx. 15KVA capacity"	Suitable UPS should be supplied with 1hrs backup approx. 10KVA capacity
BOQ(2)	Coupling Jelly 100 Litres	Coupling Jelly 100 Litres - If Required for system
BOQ (4)	Lead apron with thyroid shield -10 nos	Lead apron and Thyroid shield - 10 nos Each
BOQ(6)	ECG Electrodes 100 Nos	Deleted

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**