

**Amendment No. 4****Date: 03/09/2019****Sub: Amendment to the Tender Enquiry Document.****Ref: Tender Enquiry No.: HITES/PCD/NIB-02/19-20 dated 17.06.2019.**

The following changes are being incorporated in the referred tender enquiry for event no. 3000004111 (HPLC System) and 3000004112 (UPLC).

**SECTION I****NOTICE INVITING TENDERS (NIT)****FOR:**

<b>Sl. No.</b>	<b>Description</b>	<b>Schedule</b>
b	Closing date & time for submission of tender fee and <b>EMD in physical form</b>	<b>04.09.2019, 1400 Hrs. (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
c	Closing date & time for <b>submission of online bids</b>	<b>04.09.2019 at 1300 hrs. (IST)</b>
d	Time, date of e-tender <b>opening of online bids</b>	<b>04.09.2019, 1430 Hrs.(IST)</b>

**READ AS**

<b>Sl. No.</b>	<b>Description</b>	<b>Schedule</b>
b	Closing date & time for submission of tender fee and <b>EMD in physical form</b>	<b>17.09.2019, 1400 Hrs. (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
c	Closing date & time for <b>submission of online bids</b>	<b>17.09.2019 at 1300 hrs. (IST)</b>
d	Time, date of e-tender <b>opening of online bids</b>	<b>17.09.2019, 1430 Hrs.(IST)</b>

**Section – VII****Technical Specifications**

<b>Item no.3 HPLC System</b>		
<b>Tender Page No. &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
<b>2</b>	<b>Solvent System &amp; Pump (Quaternary Gradient):</b>	<b>2. Solvent System &amp; Pump (Quaternary Gradient):</b>
Pg 53 Para 2	Dwell volume (Total system) : ≤1ml	<b>Dwell volume (Total system) : ≤ 1.2ml</b>

All other contents of the tender enquiry including terms & conditions remain unaltered.

**Note: Prospective Bidders are also advised to check the website regularly prior to the Closing date and time of online submission of bids.**