

**M/s. HLL Infra Tech Services Ltd. (HITES)**  
**(Subsidiary of HLL Infra tech Services Ltd., A Government of India Enterprise)**  
**For & on behalf of**  
**U. N. Mehta Institute of cardiology & Research Centre,**  
**(Affiliated to B. J. Medical College, Ahmedabad, Gujarat)**

---

**CORRIGENDUM – 01(CLARIFICATION)**

**Date: 02.03.2019**

Name of the Tender: e-TENDER FOR PROCUREMENT OF SUPPLY AND INSTALLATION OF MEDICAL FURNITURE FOR NEW CONSTRUCTION OF PHASE-II, PAEDIATRIC CARDIAC HOSPITAL BUILDING (GRIHA 3 STAR RATED) BY HLL INFRATECH SERVICES LTD. ON BEHALF OF U. N. MEHTA INSTITUTE OF CARDIOLOGY & RESEARCH CENTRE, AHMEDABAD, GUJARAT

Reference: e-TENDER ID # 347990

The Corrigendum – 01 (clarification) shall be treated as part of e-tender. Scanned copy of the same to be uploaded online along with the other e-tender documents and print out of uploaded documents to be submitted in the form of hard copy along with submission of e-tender, which should be duly signed and stamped:

SN	Bidder's Queries		Clarification
	Reference to Tender Clause	Query	
1	Section VII: Technical Specification: Item no: 1 (Drug trolley) & Item no 2 (Anaesthesia trolley)	Both the trolley purpose are different need some addition and deletion in the specification	Terms and Conditions of e-Tender prevails
2	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-4	To be Read as : with convenient and smooth slide in the side out motion ( Drawers should have provision of 30 to 40 separate compartments - in above six to eight rows) Justification:	Please refer Corrigendum-01 (Addendum): Sr. no.: 1

e-TENDER FOR PROCUREMENT OF SUPPLY AND INSTALLATION OF MEDICAL FURNITURE

SN	Bidder's Queries		Clarification
	Reference to Tender Clause	Query	
		we request you to kindly correct the specification to enable other venders to participate	
3	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-5	<b>To be Read as :</b> Delete <b>Justification:</b> There is no requirement of half covered drawers we request you to kindly deleted the same	Please refer Corrigendum-01 (Addendum): Sr. no.: 2
4	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-6	<b>To be Read as :</b> Delete <b>Justification:</b> There is no requirement of half covered drawers we request you to kindly deleted the same	Please refer Corrigendum-01 (Addendum): Sr. no.: 3
5	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-7	<b>To be Read as :</b> Mounted on four 100mm castors ( two or more with brake locking arrangement) <b>Justification:</b> we request you to kindly correct the specification for the better locking of cart	Better specification is acceptable
6	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley)	<b>To be Read as :</b> should have an over bridge with bins and shelves <b>Justification:</b> We request you to kindly add over bridge with bins to keep consumables.	Terms and Conditions of e-Tender prevails
7	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-2	<b>To be Read as :</b> Flat top ABS/SS steel alloy with removable trays. <b>Justification:</b> we request you to kindly dilute the specification	Please refer Corrigendum-01 (Addendum): Sr. no.: 4
8	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-4	<b>To be Read as :</b> The front of the each drawer should contain transparent slot on which removable medicine label can be pasted	Please refer Corrigendum-01 (Addendum): Sr. no.: 5

e-TENDER FOR PROCUREMENT OF SUPPLY AND INSTALLATION OF MEDICAL FURNITURE

SN	Bidder's Queries		Clarification
	Reference to Tender Clause	Query	
		<p><b>Justification:</b> We Kindly explain the purpose of half covered drawer as our outer label will give us complete details about the content inside. However there is no requirement of half covered bins for safety &amp; security of drugs.</p>	
9	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-5	<p><b>To be Read as :</b> Mounted on four 100mm or above castors (Front two with brake locking arrangement) <b>Justification:</b> We request you to kindly dilute the specification to enable other vender to participate.</p>	Better specification is acceptable
10	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-1	<p><b>To be Read as :</b> Size: 850-1050mm L X 500- 600mm W X 1050-1200 mm H. <b>Justification:</b> We request you to kindly dilute the specification to enable other vender to participate.</p>	Please refer Corrigendum-01 (Addendum): Sr. no.: 6
11	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-2	<p><b>To be Read as :</b> Three Shelf at approx. 450mm C/c made from approx. 15mm square. <b>Justification:</b> We request you to kindly generalize the specification to enable other vender to participate.</p>	Terms and Conditions of e-Tender prevails
12	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-3	<p><b>To be Read as :</b> All materials of Stainless Steel 304 grade/ Chrome finished fitted on frame Mounted on 125mm Dia heavy duty wheel fixed on 4mm thick base plate with nut bolts and overlap 75mmL of oversize square pipe for wheel fixing. <b>Justification:</b> We request you to kindly generalize the specification to enable other vender to participate.</p>	Terms and Conditions of e-Tender prevails

e-TENDER FOR PROCUREMENT OF SUPPLY AND INSTALLATION OF MEDICAL FURNITURE

SN	Bidder's Queries		Clarification
	Reference to Tender Clause	Query	
13	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-4	<p><b>To be Read as :</b> Tray shall be fixed by 12 Nos. of SS/ABS bracket supports at all corner of frame and two heavy supports below tray frame plate.</p> <p><b>Justification:</b> We request you to kindly generalize the specification to enable other vender to participate.</p>	Terms and Conditions of e-Tender prevails
14	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-5	<p><b>To be Read as :</b> ss/ Chrome finished handle of approx. 25mm dia pipe on both side for pull/ push purpose</p> <p><b>Justification:</b> We request you to kindly generalize the specification to enable other vender to participate.</p>	Terms and Conditions of e-Tender prevails
15	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-6	<p><b>To be Read as :</b> All finish chrome finished/ matt polished</p> <p><b>Justification:</b> we request you to kindly dilute the specification for better finish</p>	Terms and Conditions of e-Tender prevails
16	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley)	<p><b>To be Read as :</b> Trolley should be ss/Chrome finished with wire shelving</p> <p><b>Justification:</b> we request you to kindly allow the chrome finished trolley for better finish and strong trolley</p>	Terms and Conditions of e-Tender prevails

\*\*\*\*\* End of Corrigendum – 01(Clarification) \*\*\*\*\*