

Corrigendum 02 in GeM portal

Bid Number: GEM/2020/B/781727 Dated: 08-09-2020

Item Category: Ultra low temperature laboratory deep freezer

1. Technical Specification: Performance Parameters: Energy consumption rate (@KWh/day) -- Point will not be considered for evaluation purpose.
2. Technical Specification: Performance Parameters: Warranty of compressor in years: Minimum 5 years.
3. Technical Specification: Additional Parameters: One 5KVA servo stabilizer to be supplied with freezer : Yes

Prebid-Clarification

Bid No:-GEM/2020/B/781727 Dated: 08-09-2020 Item Name:- Ultra Low Temperature Laboratory Deep Freezer

Date of Pre-bid Meeting - (10-09-2020 15:00:00)

Sl.No.	Tender Specification (Clause No & Page No)	Name of Bidder	Representations Received from the bidders	Pre-Bid Clarification
1	Performance Parameters: Panel reference Value:-Vacuum polyurethane foam panels (Page Number 4 of Bid Document)	Biolinx Labsystems Pvt. Ltd.	Is PUF Panel (polyurethane foam insulation) exceptable instead of mentioned VIP Panel (Vacuume Insulated Panels)?	Tender Specification shall prevail
2	Performance Parameters: Capacity in liters reference Value:-451-500 L Bid Value:-451-500 L (Page 3 of Bid Document)		It should be changed to : 400 – 500 liters.	Tender Specification shall prevail
3	Performance Parameters: Refrigerants reference Value:- CFC-free, HCFC-free non flammable refrigerants.(Page 3 of Bid Document)		It should be changed to Refrigerants : Hydrocarbon natural refrigerants CFC, HCFC free as per safety compliant EN60335-2-24.	Tender Specification shall prevail
4	Performance Parameters:- Control system reference Value:-Fully programmable microprocessor controlled with membrane keypad and eye level control panel(Page 3 of Bid Document)	Labindia	It should be changed to - Control system : Fully programmable microprocessor controlled with membrane keypad / Fully programmable microprocessor controlled with Touch Screen mounted at eye level control panel. As membrane keypad “word” is favourable to particular vendor, whenever touch screen control is better than membrane key pad.	Tender Specification shall prevail

Prebid-Clarification

Bid No:-GEM/2020/B/781727 Dated: 08-09-2020 Item Name:- Ultra Low Temperature Laboratory Deep Freezer

Date of Pre-bid Meeting - (10-09-2020 15:00:00)

Sl.No.	Tender Specification (Clause No & Page No)	Name of Bidder	Representations Received from the bidders	Pre-Bid Clarification
5	Performance Parameters:- Freezer should have heated air vent and easily accessible front panel filter reference Value: -Yes (Page 3 of Bid Document)	Instruments Pvt.Ltd.	It should be changed to : Freezer should have heated air vent / Perfect fit door sealing system with heated seal to prevent ice forming and easily accessible front panel filter. As heated air vent "word" is favourable to particular vendor.	Tender Specification shall prevail
6	Additional Parameters:- CO2 cylinder should be supplied with freezer for back up. (Page 4 of Bid Document) reference Value:-Liquid Nitrogen should be given for backup instead of CO ₂ cylinder(as per corrigendum -01)		It should be changed to : Freezer should be quoted with CO2 Back up system with CO2 gas cylinder for back up to maintain temperature during the cut of power.	As per corrigendum 1 - Liquid Nitrogen should be given for backup instead of CO ₂ cylinder
7	Additional Parameters:- Number of corrosion resistant racks reference Value:-2(Page 4 of Bid Document) Bid Value:-4 or more (as per corrigendum -01)		Since one freezer can accommodate minimum 15 corrosion resistant racks and you have asked for only 2 No. Kindly change the same accordingly.	Tender Specification shall prevail
8	Performance Parameters: Capacity in liters reference Value:-451-500 L Bid Value:-451-500 L (Page 3 of Bid Document)		500-550L	Tender Specification shall prevail

Prebid-Clarification

Bid No:-GEM/2020/B/781727 Dated: 08-09-2020 Item Name:- Ultra Low Temperature Laboratory Deep Freezer

Date of Pre-bid Meeting - (10-09-2020 15:00:00)

Sl.No.	Tender Specification (Clause No & Page No)	Name of Bidder	Representations Received from the bidders	Pre-Bid Clarification
9	Material of inside chamber reference Value:SS. (Page 3 of Bid Document)	Vision Diagnostic (I) Pvt Ltd.	Heat Treated Re-Enforced High strength Steel.	Tender Specification shall prevail
10	Energy consumption rate (@KWh/day) reference Value:-2 (Page 4 of Bid Document)		12	Point will not be considered for evaluation purpose. Suitable Amendment is issued.
11	Double door with locking lids reference Value:Yes (Page 4 of Bid Document)		Inner door:2-4	Double door refers to one outer door and an appropriate inner separator/door as per quoted model. This is to ensure maximum cooling retention.
12	Warranty of Freezer (except compressor) in Years (from the date of installation) reference Value:3 (Page 4 of Bid Document)		Warranty of freezer :5	Tender Specification shall prevail
13	Warranty of compressor in years (Page 4 of Bid Document) reference Value:5 years (as per corrigendum -01)		Warranty of compressor in years :5	Warranty of compressor in years : Minimum 5 Years, Suitable Amendment is issued.
14	One 5KVA servo stabilizer to be supplied with freezer reference Value:No (Page 4 of Bid Document)		Yes	One 5KVA servo stabilizer is required to be supplied with freezer. Suitable Amendment is issued.

Prebid-Clarification

Bid No:-GEM/2020/B/781727 Dated: 08-09-2020 Item Name:- Ultra Low Temperature Laboratory Deep Freezer

Date of Pre-bid Meeting - (10-09-2020 15:00:00)

Sl.No.	Tender Specification (Clause No & Page No)	Name of Bidder	Representations Received from the bidders	Pre-Bid Clarification
15	Additional Parameters:- CO2 cylinder should be supplied with freezer for back up.(Page 4 of Bid Document) reference Value:-Liquid Nitrogen should be given for backup instead of CO ₂ cylinder(as per corrigendum -01)		CO2 cylinder should be supplied with freezer for backup:No,as no backup is required.	As per corrigendum 1 - Liquid Nitrogen should be given for backup instead of CO ₂ cylinder
16	General Clarification			The bidder should submit catalogue/datasheet and for the same make & model of kit as offered in GeM bid.
17	General Clarification			The Bid Requirement (Allowed Values) As per GeM Category specification can be checked by browsing hyperlink in page 2 of the GeM Bidding document.