

04-05-2021

**Amendment No. 03****Sub: Amendment to the referred tender enquiry**

Ref.: Tender Enquiry Document HITES/PCD/VKNRL/TED04/2020-21 dated 23-03-2021 and subsequent amendment(s).

Following changes are being incorporated in the tender

**SECTION I****NOTICE INVITING TENDER (NIT)**

**For:**

**Tender Timeline**

Sl. No.	Description	Schedule
1	Last date and time of online submission of tender	05-05-2021, 01:00 PM
2	Date and time of tender Opening	06-05-2021, 02:00 PM
3	Last date and time of physical submission of Tender processing Fee, any other document specified in the Bidding Document	11-05-2021, 02:30 PM

**Read As:**

Sl. No.	Description	Schedule
1	Last date and time of online submission of tender	<b>19-05-2021, 01:00 PM</b>
2	Date and time of tender Opening	<b>20-05-2021, 02:00 PM</b>
3	Last date and time of physical submission of Tender processing Fee, any other document specified in the Bidding Document	<b>25-05-2021, 02:30 PM</b>

**SECTION VII****TECHNICAL SPECIFICATIONS**

<b>Item No. 01 16 Slice Whole Body CT Scanner Machine</b>		
<b>Tender Pg. &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
Pg 41 Para M.4	The advance vessel analysis, auto bone removal, small volume quantification, lung nodule assessment. CT whole body perfusion along with brain perfusion, Dental CT. (Dental reformat Package). Bolus tracking, Ct. reporting, SAS (Spiral auto start), CT fluoroscopy (Continuous CT Biopsy), paediatric protocols, Urology Package, Bone mineral Density etc. should be possible even on independent workstation.	The advance vessel analysis, auto bone removal, small volume quantification, lung nodule assessment, CT reporting, Urology Package, etc. should be possible even on independent workstation. Neuro Perfusion, Bolus tracking, SAS (Spiral auto start), Dental CT (Dental reformat Package). Paediatric protocols should be available <b>on the console/independent work station.</b>

**Note:**

- i. All other contents of the tender enquiry including terms & conditions remain unaltered.
- ii. TC compliance sheet v1.1 is being uploaded on CPP Portal for necessary actions of interested Bidders. Any other version uploaded earlier should be ignored.
- iii. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.