

M/s. HLL Infra Tech Services Ltd. (HITES)
(Subsidiary of HLL Infra tech Services Ltd., A Government of India Enterprise)
For & on behalf of
U. N. Mehta Institute of cardiology & Research Centre,
(Affiliated to B. J. Medical College, Ahmedabad, Gujarat)

CORRIGENDUM – 01 (ADDENDUM)

Date: 02.03.2019

Name of the Tender: e-TENDER FOR PROCUREMENT OF SUPPLY AND INSTALLATION OF MEDICAL FURNITURE FOR NEW CONSTRUCTION OF PHASE-II, PAEDIATRIC CARDIAC HOSPITAL BUILDING (GRIHA 3 STAR RATED) BY HLL INFRATECH SERVICES LTD. ON BEHALF OF U. N. MEHTA INSTITUTE OF CARDIOLOGY & RESEARCH CENTRE, AHMEDABAD, GUJARAT

Reference: e-TENDER ID # 347990

The Corrigendum – 01(Addendum) shall be treated as part of e-tender. Scanned copy of the same to be uploaded online along with the other e-tender documents and print out of uploaded documents to be submitted in the form of hard copy along with submission of e-tender, which should be duly signed and stamped:

Amendment to existing e-tender terms & conditions:

SN	Reference to E-Tender Clause	Amended Clause (New Clause Added)
1	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-4: With convenient and smooth slide in the slide out motion (At least 30 to 40 separate drawers – in about six to eight rows)	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-4: With convenient and smooth slide in the side out motion. a. There should be 6 to 8 rows of Drawers. b. These 6 to 8 rows of drawers should have provision of total 30 to 40 separate compartment.
2	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-5: The front of the each drawer should be half covered on which removable medicine label	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-5: The front of the each drawer should have facility on which removable medicine label can be pasted

e-TENDER FOR PROCUREMENT OF SUPPLY AND INSTALLATION OF MEDICAL FURNITURE

SN	Reference to E-Tender Clause	Amended Clause (New Clause Added)
3	Section-VII: Technical Specification: Item no 2 (Anaesthesia trolley): Point-6: Can be pasted and upper half open to see the contents inside.	Deleted
4	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-2: Flat top ABS with removable trays.	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-2: Flat Top ABS/SS steel alloy with removable trays
5	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-4: The front of the each drawer should be half covered on which removable medicine label Can be pasted and upper half open to see the contents inside.	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-4: The front of the each drawer should have facility on which removable medicine label can be pasted
6	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-1: Size: 1050mm L X 600mm W X 1050mm H	Section-VII: Technical Specification: Item no: 3 (Material Transfer trolley): Point-1: Size: 900-1050 mm L X 550-600 mm W X 950-1050 mm H
7.	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-3: Should have multiple long drawers to hold drug strips made of high quality ABS With convenient and smooth slide in the slide out motion (At least 30 to 40 separate drawers – in about six to eight rows)	Section-VII: Technical Specification: Item no: 1 (Drug trolley): Point-3: Should have multiple long drawers to hold drug strips made of high quality ABS With convenient and smooth slide in the slide out motion. a. There should be 6 to 8 rows of Drawers. b. These 6 to 8 rows of drawers should have provision of total 30 to 40 separate compartment.
7	<p>SECTION I</p> <p>NOTICE INVITING TENDERS (NIT)</p> <p>(A) E tender Enquiry No.: HITES/PCD/UNM-AHD/05/MED-FURN-RT/18-19 Dated: 06.02.2019</p> <p>Important Dates and Venue for Events:</p>	<p>SECTION I</p> <p>NOTICE INVITING TENDERS (NIT)</p> <p>(A) E tender Enquiry No.: HITES/PCD/UNM-AHD/05/MED-FURN-RT/18-19 Dated: 06.02.2019</p> <p>Important Dates and Venue for Events:</p>

e-TENDER FOR PROCUREMENT OF SUPPLY AND INSTALLATION OF MEDICAL FURNITURE

SN	Reference to E-Tender Clause			Amended Clause (New Clause Added)		
	SI. No.	Description	Schedule	SI. No.	Description	Schedule
	i.	Pre-bid Meeting Date & Time	13.02.2019, 12.30 P.M (Afternoon)	i.	Pre-bid Meeting Date & Time	13.02.2019, 12.30 P.M (Afternoon)
	ii.	Pre-bid Meeting Venue	4 th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.	ii.	Pre-bid Meeting Venue	4 th Floor, Central Lecture Theatre, U. N. Mehta Institute of Cardiology & Research Centre, (Affiliated to B. J. Medical College), Asarwa, Ahmedabad – 380016 Gujarat, India.
	iii.	Due date & time for submission of e tender in e tender Portal	06.03.2019, 1.00 P.M (After Noon)	iii.	Due date & time for submission of e tender in e tender Portal	19.03.2019, 1.00 P.M (After Noon)
	iv.	Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form	06.03.2019, 2.30 P.M (After Noon)	iv.	Closing date & time for submission of E tender Processing Fee, EMD and Technical bid in physical form	19.03.2019, 2.30 P.M (After Noon)
	v.	Date and Time of opening of Techno - Commercial e tenders	06.03.2019, 3.00 P.M (After noon)	v.	Date and Time of opening of Techno - Commercial e tenders	19.03.2019, 3.00 P.M (After noon)
	vi.	Venue for opening of E tenders	M/s. HLL Infra Tech Services Ltd., B14-A, Sector-62, Noida (UP) – 201307 India.	vi.	Venue for opening of E tenders	M/s. HLL Infra Tech Services Ltd., B14-A, Sector-62, Noida (UP) – 201307 India.

***** End of Corrigendum – 01 (Addendum) *****