

Corrigendum 02 in GeM portal
Bid Number: GEM/2020/B/881345 dated 19-11-2020
Item Name: Solar direct drive combined refrigerator/freezer (small)

1: Interested bidders are required to sign "Bid Security Declaration" accepting that if they withdraw or modify their Bids during the period of validity, or if they are awarded the contract and they fail to sign the contract, or to submit a performance security before the deadline defined in the request for bids document, they will be suspended for the period of 6 months from being eligible to submit Bids for contracts with the entity that invited the Bids.

2: Existing

One Bid per Bidder: A firm shall submit only one bid either individually or as a partner of a joint venture. A firm that submits either individually or, as a member of a joint venture, more than one bid will cause all the proposals with the firms participation to be disqualified.

Read As

One Bid per Bidder: An OEM can either participate directly or can only authorize one bidder to quote on their behalf. In case of submission of multiple offers by an OEM, directly or through its authorized agent(s), all such offers are liable to be rejected.

3: Existing: Provision for drainage plug for the waste water in both the compartment

Read as: Provision for drainage plug for the waste water in both the compartment. **In case of vertical type, provision of drainage has to be provided in vaccine chamber.**

4: Existing: Have legs in the base with rotating screw type height adjustments to balance the weight on uneven floor

Read as: Appropriate mechanism shall be there, maintaining levelled position for equipment

5: Existing: Tubing - Type compression cycled, (both for refrigeration & insulation all system tubing, suction tube, freezer tube, condensing tube) should be min 99.97% of pure copper coil

Read as: Type compression cycled, (both for refrigeration & insulation all system tubing, suction tube, freezer tube, condensing tube) **should be pure copper coil**

6: Existing: Climate zone - Hot (Design and tested at temperature 5 to 43 deg C without solar energy

Read as: Climate zone - Hot (Design and tested at temperature **+10 to +43 deg C** without solar energy

7: Existing: Net vaccine storage capacity of refrigerator compartment-**58 L**

Read as: Net vaccine storage capacity of refrigerator compartment- **35-70 L**

8: Existing: Temperature of compartment - Refrigerated compartment temperature between 2 to 8 deg C and Freezing compartment temperature between -25 to -15 degree Celsius

Read as: Temperature of compartment - Refrigerated compartment temperature between +2 to +8 deg C and Freezing compartment temperature **should be in the range of 0° C to -20° C**

9: Existing: Material of construction - Roto moulded

Read as: Material of construction - **Roto moulded or Stainless steel (SS304) rusts free**

10: Existing: Control unit should be provided for setting of temperature and display with following feature : Separate digital display of cabinet temperature required, Power on LCD, LED indicator, **Audible alarm and visual alarm for both unit separately**, Unit should be sealed, protected from dust moisture or condensed water falling over it, Sensor accuracy of digital controller +-0.5 degree Celsius

Read as: Control unit should be provided for setting of temperature and display with following feature : Separate digital display of cabinet temperature required, Power on LCD, LED indicator, **Audible alarm and visual alarm for both unit separately (optional)**, Unit should be sealed, protected from dust moisture or condensed water falling over it, Sensor accuracy of digital controller +-0.5 degree Celsius

11: **Existing:** Accessories spares & Warranty - Warranty (min) Comprehensive maintenance of 10 year after installation

Read as: Accessories spares & Warranty - Warranty (min): **(after installation) for the 5 years period**

12: **Existing:** Accessories spares & Warranty - Accessories spares Vaccine: Vaccine storage basket allowing free circulation of air, having the size to be able to accommodate in the unit and suitable to match the net volume requirement

Read as: Accessories spares & Warranty - Accessories spares Vaccine: **Vaccine storage basket/wire shelves** allowing free circulation of air, having the size to be able to accommodate in the unit and suitable to match the net volume requirement.

13: Existing: Duration of ePBG required (Months): 122

Read As: Duration of ePBG required (Months): 62