

14-10-2020

Amendment No. 9**Sub: Amendment to the referred tender enquiry****Ref.: HITES/PCD/AIIMS-IV/25/OBG/19-20 dated 29-01-2020 read with Notice on Extension of Due Dates dated 25.03.2020 and amendment no. 1, 2, 3, 4, 5, 6 & 7 dated 09.03.2020, 11.03.2020, 11.08.2020, 19.08.2020, 03.09.2020, 16.09.2020, 28.09.2020 33& 14.10.2020 respectively.****The following changes are being incorporated in the following mentioned Tender ID only**

- 2020_HLL_41504_14 (Laparoscopic Surgery Set with Hysterescope & resectoscope with High Definition Camera & Monitor).
- 2020_HLL_41504_15 (Portable Ultrasound & Colour Doppler)
- 2020_HLL_41504_17 (Gynae OT Table)
- 2020_HLL_41504_22 (Hysterectomy set)
- 2020_HLL_41504_23 (MTP Suction)

SECTION I**NOTICE INVITING e-TENDER (e-NIT)****2) Tender Timeline:****Existing:**

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	28-10-2020, 1200 hrs
d.	Closing date & time for submission of tender processing fee and EMD in physical form*	29-10-2020, 1400 hrs
e.	Time and date of opening of online bids	29-10-2020, 1430 hrs

Read as:

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	11-11-2020, 1200 hrs
d.	Closing date & time for submission of tender processing fee and EMD in physical form*	12-11-2020, 1400 hrs
e.	Time and date of opening of online bids	12-11-2020, 1430 hrs

SECTION VII
Technical Specification

Item No. 1 Laparoscopic Surgery Set with Hysterescope & resectoscope with High Definition Camera & Monitor (2020_HLL_41504_1)		
Tender Page & Para	TENDER SPECIFICATION	AMENDED AS
Pg 45 Para 2. 1.	Telescopes, Insufflator, Camera Control Unit and Camera head, HD Medical grade monitor, Light source, Image management system, should be from same manufacturer with USFDA or European CE approved Hysteroscope & Resectoscope should be from same principal manufacturer; incase these items are not manufactured by the principal manufacturer, the same shall be offered from a second manufacturer (In this case the principal manufacturer should submit an undertaking for non production of these items with USFDA or European CE approved).	Telescopes, Insufflator, Suction irrigation Unit, Camera Control Unit and Camera head, HD Medical grade monitor, Light source, Image management system, Hysteroscope & Resectoscope should be from same manufacturer with USFDA or European CE or BIS approved
Pg 45 Para 2. 3.	Regarding hand instruments : 70% - 80% of the hand instruments should be supplied from same principal manufacturer and all the instruments should be USFDA or European CE approved	Regarding hand instruments : 80% of the hand instruments should be supplied from same principal manufacturer and all the instruments should be USFDA or European CE or BIS approved
Pg 49 Para 3.11 a)	LED Light Source equivalent to 300 Watts	LED Light Source, having illuminance equivalent to 300 Watts Xenon lamp
Pg 50 Para 3.13 k)	Full HD recording, Medical grade computer and Monitor, Touch screen, Minimum 1 TB storage memory. It should have window based operating system, minimum Windows - XP.	Full HD recording , Medical grade Monitor , Touch screen, Minimum 1 TB storage memory. It should have window based operating system.

Item No. 2 Portable Ultrasound & Colour Doppler (2020_HLL_41504_2)		
Tender Page & Para	TENDER SPECIFICATION	AMENDED AS
Pg 54 Para 8	Transducers (one each):	Transducers (one each): Added para : Variation of (± 1 MHz) in the limits specified for the operating frequency for all transducer frequency is acceptable .

Item No. 4 Gynae OT Table (2020_HLL_41504_4)		
Tender Page & Para	TENDER SPECIFICATION	AMENDED AS
Pg 58 First Para	Multipurpose electro hydraulic with manual override mobile Table with divided leg section suitable for all Gynaecological surgical procedures, complete with 5 cm mattress and corded handset.	Multipurpose electro hydraulic with manual/ electrtical override mobile Table with divided leg section suitable for all Gynaecological surgical procedures, complete with 5 cm mattress and corded handset.
Pg 58 First Para 16	4 or 5 sections tabletop, which should be made of a special scratch resistant, hardwearing and easy to clean material. Base column cover to be made of 100% stainless steel alloy and stainless steel.	4 or 5 sections tabletop, which should be made of a special scratch resistant, hardwearing and easy to clean material. Base and entire column cover to be made of 100% stainless steel alloy and stainless steel.
Pg 58 Para 21.	The robust handset should offer 8 controls namely Trend/ Reverse Trend, Lateral Tilt, Flexion/ Extension and Height functions.	The robust handset should offer minimum 8 controls namely Trendelenburg and Reverse Trendelenburg , Lateral Tilt, Zero position, Longitudinal slide and Height functions.
Pg 59 Para F	Should be USFDA or European CE or ISO 13485 along with EC Declaration of Conformity or BIS approved product	Should be USFDA or European CE or ISO 13485 (issued by a notified body) along with EC Declaration of Conformity or BIS approved product

Item No. 10 MTP Suction (2020_HLL_41504_10)		
Tender Page & Para	TENDER SPECIFICATION	AMENDED AS
Pg 66 Para 8	Equipment should be manufactured by well-known International Company	Deleted

All other contents of the Tender enquiry including terms & conditions remain unaltered.

Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**
- II. The earnest money shall be valid for a period of forty-five (45) days beyond the validity period of the tender. As validity period of Tender as per Clause 20 of GIT is 180 days, the EMD shall be valid for 225 days from Techno – Commercial Tender opening date.**
- III. Technical Compliance Sheet ‘Ver 1.1’ and Price bid Sheet ‘Ver 1.1’ in excel form against referred tender are being uploaded on CPPP. The same need to be downloaded for submission.**