

**Amendment No. 3****Sub: Amendment No. 1 to the TED.****24-07-2018****Ref: HITES/PCD/MP/03/PRECLINICAL-RT-01/18-19 Dated: 08.06.2018 .**

The following amendments are being incorporated to the TED :

**SECTION – I  
NOTICE INVITING TENDER (NIT)****Existing for Sch. 8, 3000002881, Microscope Trinocular (Teacher),:**

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	25.07.2018, 12:00 PM
d.	Closing date & time for submission of tender processing fee and EMD in physical form*	25.07.2018, 02:00 PM
e.	Time and date of opening of online bids	25.07.2018, 02:30 PM

**Read As (Sch. 8, 3000002881, Microscope Trinocular (Teacher),):**

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	<b>07.08.2018, 12:00 PM</b>
d.	Closing date & time for submission of <b>tender processing fee and EMD in physical form*</b>	<b>07.08.2018, 02:00 PM</b>
e.	Time and date of opening of online bids	<b>07.08.2018, 02:30 PM</b>

**Section – VII  
Technical Specifications**

<b>Schedule No: 08:-Trinocular RESEARCH MICROSCOPE</b>		
<b>Para</b>	<b>Existing Specification</b>	<b>Read As</b>
6	Condenser system 0.90 Dry and 1.25 Oil & (Dark field slider ,Polarization attachment [ Polarizer and analyser]	Universal condenser for bright field, Dark field, phase contrast & Polarization application attachment [ Polarizer and analyser]
29	OPTIONAL ( price should be quoted separately Price will be considered for ranking purpose) - Fluorescence attachment with HG 100W ( 10 nos with system)/LED Fluorescent light source & three fluorescence filter to cover DAPI/Hoechest , FITC /GFP ,TRITC /Rhodamine along with suitable 3 to 5 MP color camera capable of Bright field & immuno fluorescence application.	OPTIONAL ( price should be quoted separately Price will be considered for ranking purpose) - Fluorescence attachment with HG 100W /LED Fluorescent light source & three fluorescence filter to cover DAPI/Hoechest , FITC /GFP ,TRITC /Rhodamine along with suitable 3 to 5 MP color camera capable of Bright field & immuno fluorescence application.
	Added para:- Sextuple Nosepiece to accomodate 6 Objectives.	<b>Quintuple/ Sextuple Nosepiece to accomodate 5 to 6 Objectives with Analyser slot to polarizing microscopy application.</b>

**Note:**

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.**
- ii. **If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening.**

**All other contents of the tender enquiry including terms & conditions remain unaltered.**