

19.09.2022

Amendment No. 7

Sub: Technical Amendment to the referred tender enquiry

Ref.: HITES/PCD/AIIMS-IV/53/CSSD/22-23 dated. 01-06-2022 Read with its Amendment no 1, 2, 3, 4, 5, 6 and clarification no 1 dated: 12.07.2022, 26.07.2022, 28.07.2022, 12.08.2022 08.08.2022, 05.09.2022 and 22.08.2022 respectively.

The following changes are being incorporated in the above referred Tender Enquiry Document.

**SECTION I
NOTICE INVITING TENDER (NIT)**

2) Tender timeline:

Existing:

Sl. No.	Description	Schedule
b.	Closing date & time for submission of online bids	19-09-2022, 1400 hrs
c.	Closing date & time for submission of tender processing fee in physical form*	20-09-2022, 1400 hrs
d.	Time and date of opening of online bids	20-09-2022, 1430 hrs

Read as:

Sl. No.	Description	Schedule
b.	Closing date & time for submission of online bids	28-09-2022, 1400 hrs
c.	Closing date & time for submission of tender processing fee in physical form*	29-09-2022, 1400 hrs
d.	Time and date of opening of online bids	29-09-2022, 1430 hrs

SECTION VII
Technical Specification

Tender Page & Para	TENDER SPECIFICATION	AMENDED AS
II. 1	Washer disinfecter (300 to 350L) with accessories	
Page 46	Should be US FDA/European CE certified with four digit notified body number.	Should be BIS / US FDA/European CE certified with four digit notified body number.
II. 4	Ultrasonic Cleaner (40 L)	
Page 47; Para h	Ultrasonic cleaner should be either ISO 13485 with Declaration of conformity as per European CE norms or US FDA certified.	Ultrasonic cleaner should be either BIS or ISO 13485 with Declaration of conformity as per European CE norms or US FDA certified.
VII. 1	Horizontal Sterilizer 500 litre or more with minimum 8 STU capacity and Accessories	
Page 54 Para n	It should meet EN ISO / IEC directives and product should be US FDA/European CE certified with four digit notified body number.	It should meet EN ISO / IEC directives and product should be BIS / US FDA/European CE certified with four digit notified body number.
VII. 2	Sterilizer 250 L or more and with minimum 4 STU capacity with Accessories:	
Page 57 Para n.1	It should meet EN ISO / IEC directives and product should be US FDA/European CE certified with four digit notified body number.	It should meet EN ISO / IEC directives and product should be BIS / US FDA/European CE certified with four digit notified body number.
VII. 3	Table Top Sterilizer with Accessories	
Page 57; Para 5	Quality Standards: Sterilizer should be US FDA/European CE certified with four digit notified body number	Quality Standards: Sterilizer should be BIS / US FDA/European CE certified with four digit notified body number
VII. 6	Heat Sealing Machine	
Page 59; Para m	Rotary heat sealer should be European CE /US FDA certified.	Rotary heat sealer should be BIS / European CE /US FDA certified.
XV. Turn Key Works		
Page 69 Para 11	One Point Water supply, Drain & Electricity shall be provided by respective Institute. Rest of the works to be executed by the bidder	One Point Water supply, Drain & Electricity shall be provided by respective Institute. Chilled & hot water supply for Air conditioning and fresh air supply duct in the AHU room shall be provided by the Institute at designated points of AHU near CSSD. The Dedicated drain for AHU room shall also be given by institute. Rest of the works shall be executed by the bidder.

All other contents of the Tender enquiry including terms & conditions remain unaltered.
Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.**
- II. If EMD is submitted in the form of BG/FDR, then the validity of the BG/FDR should be at least 225 days from the date of tender opening.**