# Amendment No. 07

Date: 19-02-2024

Sub: Amendment No. 07 to the Tender Enquiry Document

Ref: Tender No: HITES/PCD/AIIMS-IV/62/RT/Rad-Onco/23-24/ dated 05-12-2023

Tender ID: 2023\_HITE\_176706\_1

### <u>Section I</u> Notice Inviting Tenders (NIT)

#### **Tender timeline:**

#### For

Sl. No.	Description	Schedule
4.	Closing date & time for submission of online bids	<b>26-02-2024</b> , 01:00 PM
5.	Closing date & time for submission of <b>tender processing fee and EMD in physical form*</b>	<b>26-02-2024,</b> 02:00 PM
6.	Time and date of opening of online bids	<b>26-02-2024</b> , 02:30 PM
7.	Venue for opening of online bids and submission of tender processing fee and EMD	HLL Infra Tech Services Limited, Procurement & Consultancy Services Division, B-14 A, Sector-62, Noida-201307

#### Read As

Sl. No.	Description	Schedule
4.	Closing date & time for submission of online bids	<b>01-03-2024</b> , 01:00 PM
5.	Closing date & time for submission of <b>tender processing fee and EMD in physical form*</b>	<b>01-03-2024,</b> 02:00 PM
6.	Time and date of opening of online bids	<b>01-03-2024</b> , 02:30 PM
7.	Venue for opening of online bids and submission of tender processing fee and EMD	HLL Infra Tech Services Limited, Procurement & Consultancy Services Division, B-14 A, Sector-62, Noida-201307

## Section VII Technical Specifications

Para No	Tender Technical Specification	Amended as
7	Added Para : Technical Specification for Surface Guided Radiation Therapy (SGRT) System:	
	The Vender to provide Optical Surface Imaging system for Surface guided Radiation Therapy available in the market for patient setup, Positioning, Intra Fraction motion monitoring for all cancer treatments, including but not limited to DIBH, Pelvis, SRS, SBRT, Paediatrics and at non-coplanar gantry angles with following specifications:	
	9. System should be able to capture multiple images within a specific time frame to obtain mid-breathing reference image to assist free breathing patient treatment.	Deleted
	NA	Added Para 9 under technical specifications of SGRT System:  9) The offered system should be capable for both adult and pediatric treatment.

All other terms and conditions of the tender enquiry remain unaltered