

Response To Pre-Bid Queries (Pre-Bid date: 04.10.2022)				
Bid Number: GEM/2022/B/2583967 Dated: 28-09-2022				
Item Name: Defibrillator with paediatric pads with AED Facility				
Sr. No.	ATC DOCUMENT PARA AND PAGE	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	NA	Suggested by bidder	Request you to add feature of Cardiac Monitor with defibrillator which should be upgradable for SpO2, NIBP, Temperature and Etco2 for wider participants.	Added Para: Feature of Cardiac Monitor with defibrillator which should be upgradable for SpO2, NIBP, Temperature and Etco2 for wider participants.
2	Para 1 Point 7	The machine should be able to defibrillate Adult and pediatric patients. Also 2 in 1 Or separate paddles for pediatric patients should be supplied (one set with each unit).	as per prebd discussion	Amended as: The machine should be able to defibrillate Adult and pediatric patients. Also 2 in 1 Or separate paddles for pediatric patients should be supplied.
3	NA	Suggested by bidder	Required Disposable pads and accessories need to be mentioned along with configuration of the equipment for the proper clarity.	Accessories should be supplied with each unit. 1. ECG cable with accessories-1 Set 2. Disposable pad- 1 No 3. ECG rolls 10 Nos 4. Adult paddles-1No 5. Paediatric Paddles- 2 Nos
4	NA	Suggested by bidder	Machine should be supplied with good Quality trolley having drawer for keeping Accessories. Request to amend this with the same for maintaining the product quality.	No change considered
5	Para 10 Page 01	It should have ability, by inbuilt/integrated structure to measure chest compression rate and depth in real time with both visual & audible feedback and optional CPR index on screen.	This is a patent feature of ZOLL Medical. It is requested to please delete the same.	No change considered
			Delete this point Justifications: for wider participation	
6	Para 15 Page 01	Should have internal memory of at least 72 hours of all data and 100 events (duration, number of shocks, date and time of shocks)	Should have internal memory of at least 24hours or more of all data Justifications: for wider participation	No change considered

Response To Pre-Bid Queries (Pre-Bid date: 04.10.2022)				
Bid Number: GEM/2022/B/2583967 Dated: 28-09-2022				
Item Name: Defibrillator with paediatric pads with AED Facility				
Sr. No.	ATC DOCUMENT PARA AND PAGE	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
7	NA	Suggested by bidder	It is requested, as now defibrillators are under import restriction category from 1st October, 2022. Please ask all the bidders to attach CDSCO Import License for the defibrillator quoted model.	Added Para:- Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model. In case the vendor has not yet obtained import/manufacturing license from CDSCO for the quoted model, proof of application for CDSCO medical device license to be submitted in the bid document and valid CDSCO license to be produced at the time of supply/ NOA for the quoted model