

Response To Pre-Bid Queries (Pre-Bid date: 04.10.2022)				
Bid Number: GEM/2022/B/2558999 Dated: 27-09-2022				
Item Name: Radiant Warmer				
Sr. No.	Buyer and specification pg and Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Para 23 Page 1	Equipment should be European CE with four digit notified body number or US FDA/BIS Certified.	as per prebid discussion	Should have BIS or US FDA or European CE with four digit notified body number certificate and certificate to be submitted. OR Should meet IEC 60601-1, IEC 60601-1-2 & IEC 60601-2-21 standards and should submit valid test report for the quoted model.