

Response To Pre-Bid Queries (Pre-Bid date: 21.12.2022)					
GeM Bid Number:GEM/2022/B/2857685 Dated: 14-12-2022					
Item Name: Portable Ventilator					
Sr. No.	Buyer uploaded ATC document Para and Page	Tender Specification	Name of the Firm	Representation received from Bidder	COMMITTEE RECOMMENDATION
1	Para 1 and Page 1	The ventilator should be a turbine operated, with in-built touch screen display of minimum 7", small in size and weight should less than 4 kg.	M/s Allied	The ventilator should be a turbine operated, with in-built touch screen display of minimum 7", small in size and weight should not be more than 7.5kg. (main unit with battery). Reason: The asked specification is limiting and restricting the competition. So, we request you to kindly amend this point for healthier and wider participation.	The ventilator should be a turbine operated, with in-built touch screen display of minimum 7", small in size and weight should less than <b>5 kg</b> .
			M/s Schiller	The ventilator should be a turbine operated, with in- built touch screen display of minimum 7", small in size and weight should less than 6 kg. Reason: Kindly allow for wider participation	
			M/s MGM Associates	Please consider weight should be 7 kg.	
2	Para 2 , point 2 and Page 1	Pressure support, Pressure control/ Assisted Pressure Control with tidal volume security- PSV, PCV, APCV + Vt security	M/s Schiller	Pressure support, Pressure control/ Assisted Pressure Control with tidal volume security- PSV, PCV, APCV + Vt security or equivalent Reason: Kindly allow for wider participation	Pressure support, Pressure control/ Assisted Pressure Control with tidal volume security- <b>PSV</b> , PCV, APCV + Vt security or equivalent
3	Para 3, page 1	Should have option for Spo2 Monitoring	M/s Schiller	Kindly remove this point Reason: Reason: Kindly allow for wider participation	No change considered
4	Para 3 and page 1	Tidal volume delivery should be from 50- 2500 ml for adult patients	M/s MGM Associates	Please consider Tidal Volume from 50-2000 ml	Tidal volume <b>50-2000 ml</b>
			M/s Schiller	Tidal volume delivery should be from 50- 2200 ml or more for adult patients. Reason: Kindly allow for wider participation	
5	Para 8 and Page 2	It should have an internal battery backup of minimum 6-8 hrs, should be extendable to 24 hours.	M/s Allied	It should have internal battery backup of minimum 6 hrs, Should be extendable to 4hr Reason: The asked specification is limiting and restricting the competition. So, we request you to kindly amend this point for healthier and wider participation.	No change considered

Response To Pre-Bid Queries (Pre-Bid date: 21.12.2022)					
GeM Bid Number:GEM/2022/B/2857685 Dated: 14-12-2022					
Item Name: Portable Ventilator					
Sr. No.	Buyer uploaded ATC document Para and Page	Tender Specification	Name of the Firm	Representation received from Bidder	COMMITTEE RECOMMENDATION
6	Page 11, page 1	Inspiratory pressure must be from 3-40 in NIV and 3-50 in invasive ventilation mode for pediatric and 5-50 cm H2O for adults.	M/s Schiller	Inspiratory pressure must be 5-50 cmH2O or more. Reason: Kindly allow for wider participation	Inspiratory pressure must be <b>5-50 cmH2O or more.</b>
7	Para 19, page 1	Should have a programming memorization of 4 preset configurations of ventilation.	M/s Schiller	Kindly remove this point Reason: Kindly allow for wider participation	No change considered
8	Para 20, page 1	Should have Display indicating, see wear and tear of turbine.	M/s Schiller	Kindly remove this point Reason: Kindly allow for wider participation	No change considered
9	Para 22, page 1	Should have the following trends for 30 days.	M/s Schiller	Should have the following trends for 72 hrs or more Reason: Kindly allow for wider participation	Should have the following trends for <b>15 days</b>
10	Para 24 and Page 2	Should comply with the US FDA and European CE /BIS international quality and safety standards	M/s Schiller	Should comply with the US FDA or European CE /BIS international quality and safety standards Reason: Kindly allow for wider participation	Equipemnt should be European CE with four digit notified body number or US FDA or BIS approved for the quoted model and certificate to be submitted.
			M/s Allied	Should comply with the USFDA /European CE and safety standards. Reason: Kindly accept European CE or equivalent Indian standards this is in line with the Govt of INDIA ruling (Office Memorandum from EPW division enclosed for reference copy enclosed). Moreover, we request you to encourage Indian companies as per the directions of our Honorable Prime Minister under make in India. We match all the specifications and have a proven track record having installations across India. We hope that you will give us a chance and amend the specifications as per Govt. The USFDA certification is available with few companies which is restricting the competition. Hence, we request you to kindly amend this point for healthier competitions as well as wider participation as wider participation	

Response To Pre-Bid Queries (Pre-Bid date: 21.12.2022)					
GeM Bid Number:GEM/2022/B/2857685 Dated: 14-12-2022					
Item Name: Portable Ventilator					
Sr. No.	Buyer uploaded ATC document Para and Page	Tender Specification	Name of the Firm	Representation received from Bidder	COMMITTEE RECOMMENDATION
11	Para 25 and Page 2	Machine should have remote monitoring facility.	M/s Allied	Kindly delete this point. Reason: Please note that the asked specification is available with few companies like Hamilton and ResMed and this particular point is restricting the competition. So, we request you to kindly delete this point for healthier and wider participation.	No change considered
			M/s Schiller	Kindly remove this point Reason: Kindly remove this point for wider participation	
12	NA	Suggestion by Bidder	M/s MGM Associates	BIS Standard ISO 10651-3:1997 standard in Transport Ventilator, which is now also Bureau of Indian Standards Lung Ventilator for Medical use – Part 3, Particular requirements for Emergency & Transport Ventilators. - EN 794-3 - EN 1789 - IEC 60601-1-12	Not Considered