

Response To Pre-Bid Queries (Pre-Bid date: 21.12.2022)					
GeM Bid Number: GEM/2022/B/2860959 Dated: 16-12-2022					
Item Name: Neonatal open care					
Sr. No.	Buyer Specification Document para and pg	Tender Specification	Name of the Firm	Representation received from Bidder	COMMITTEE RECOMMENDATION
1	Para , pg 1	Should have built in electronic weighing scale with integrated display in the control panel, minimum 0-10 kg range, Tare facility with bypass switch to allow change of linen etc without removing the baby should be provided (Optional)	NA	as per prebid discussion	Should have built in electronic weighing scale with integrated display in the control panel, minimum 0-10 kg range, Tare facility with bypass switch to allow change of linen etc without removing the baby should be provided
2	Para , pg 2	User manual with trouble shooting guidance in English.	NA	as per prebid discussion	Deleted
3	Para , pg 2	Technical manual with maintenance and first line technical intervention instructions, in English.	NA	as per prebid discussion	Deleted
4	Para , pg 2	Device is produced by ISO 13485:2003 certified manufacturer and Safety feature (govt certified)	NA	as per prebid discussion	Equipemnt should be European CE with four digit notified body number or US FDA or BIS approved for the quoted model and certificate to be submitted.