

Response To Pre-Bid Queries (Pre-Bid date: 04.10.2022)				
Global Tender Enquiry No.: HITES/PCD/MCH-AIIMS/07/22-23 Dated: 28-09-2022				
(Tender ID: 2022_HLL_130264_1) Grossing Stations				
Sr. No.	Tender Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 45 Para 4 Dimensions	Length: minimum 3 1/2 feet Depth: minimum 2 1/2 feet Height: minimum 3 feet	As per prebid discussion	Amended as: Length: 5 feet Depth: 3 feet Height: 5-7 feet
2	Pg 45 Para 9	Sink: Removable seamless deep sink with hot/cold water facility for plugging in of additional appliances on work surface.	As per prebid discussion	Amended as: Sink: Removable/fixed seamless deep sink with hot/cold water facility for plugging in of additional appliances on work surface
3	Pg 46 Para 22	Programmable microprocessor allows clock time to be changed and the filter replacement alarm to be reset.	As per prebid discussion	Amended as: Deleted