

09-06-2021

Amendment No. 4**Sub: Amendment No. 4 to the referred tender enquiry****Ref.: Tender Enquiry No. HITES/PCD/MP/CLINICAL/RC-06/20-21 dated 18.03.2021**

The following changes are being incorporated in the above referred Tender Enquiry Document

**SECTION I
NOTICE INVITING TENDER (NIT)****For:****(2) Tender Timeline:****Existing:**

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	14.06.2021, 13:00 hrs
d.	Closing date & time for submission of tender processing fee in physical form	14.06.2021, 14:00 hrs
e.	Time and date of opening of online bids	15.06.2021, 14:30 hrs

Read as:

Sl. No.	Description	Schedule
c.	Closing date & time for submission of online bids	23.06.2021, 13:00 hrs
d.	Closing date & time for submission of tender processing fee in physical form	23.06.2021, 14:00 hrs
e.	Time and date of opening of online bids	24.06.2021, 14:30 hrs

**SECTION VII
TECHNICAL SPECIFICATIONS**

Technical Amendment		
Tender Page & Para	Existing Para:-	Read as:-
Sch. No. 1 Anaesthesia Workstation (2021_HLL_73642_1)		
Page 44 10	Airway Monitoring	
a	Integrated monitor (6" or more color display/EL) for electronic monitoring and display of following set and measured values	Integrated monitor (8" or more color display/EL) for electronic monitoring and display of following set and measured values

Technical Amendment		
Tender Page & Para	Existing Para:-	Read as:-
Page 45	BOQ	
5	Note: Each Institute will be supplied 2 No. Halothane 1 desflurane Vaporizer.in addition to the standard configuration of Sevoflurane vaporizer & Isoflurane vaporizer.	Deleted
Sch. No. 4 Ventilator Neonatal with HFO/ Mechanical Ventilator with HFO (2021_HLL_73642_4)		
Page 50 8	Digital display : Should have integrated high resolution LCD screen minimum 10" or more color display with touch screen facility for real-time display of scalar (Pressure, Flow and Volume against time) and loop (Pressure-volume, volume-flow and pressure- flow). Graphic display of at least 3 waveforms together out of choice of flow, volume and pressure versus time with a facility to freeze these waveforms. Facility for loops together with a facility to freeze the same.	Digital display : Should have integrated high resolution LCD or TFT screen minimum 10" or more color display with touch screen facility for real-time display of scalar (Pressure, Flow and Volume against time) and loop (Pressure-volume, volume-flow and pressure- flow). Graphic display of at least 3 waveforms together out of choice of flow, volume and pressure versus time with a facility to freeze these waveforms. Facility for loops together with a facility to freeze the same.
Page 51 a.1	Trigger delay 30 - 60 ms	Deleted
I	HFO rate 5 to 20Hz	HFO rate 5 to 15Hz or more
SN	BOQ	
3	Medical air compressor (OEM)	Medical Air compressor should be USFDA or European CE or BIS complaint
Sch. No. 6 Transport Incubator with ventilator (2021_HLL_73642_6)		
Page 53 9	Two 10L integrated oxygen cylinders, regulator and flow meter with compatible connectors for refilling	Deleted
Sch. No. 7 Laparoscopic Surgery Set with Hysteroscope & Resectoscope with High Definition Camera & Monitor (2021_HLL_73642_7)		
Page 57 c.	Atraumatic trocar sleeve with pyramidal trocar 12mm	Atraumatic trocar sleeve with pyramidal trocar 12-15mm
3.9	CAMERA CONTROL UNIT & CAMERA HEAD	

Technical Amendment		
Tender Page & Para	Existing Para:-	Read as:-
c.	Camera control unit with three chip HD camera head having HD CCD chip of same aspect ratio of 16:9 and camera control unit should be able to produce Full HD output of minimum 3 nos	Camera control unit with three chip HD camera head having HD CCD chip/CMOS of same aspect ratio of 16:9 and camera control unit should be able to produce Full HD output of minimum 3 nos
d.	Three chip camera head should produce at head itself Pure Digital Signal with High Definition video (1920 * 1080P) with aspect ratio of CCD chip and video format of 16:9 or 16:10.	Three chip camera head should produce at head itself Pure Digital Signal with High Definition video (1920 * 1080P) with aspect ratio of CCD/CMOS chip and video format of 16:9 or 16:10.
g.	Three Chip Camera control unit should be compatible with all the tree chip camera head .	Deleted
e	Input Key Board for Character Generator, 5 pole Din	Input Key Board/ Touch screen for Character Generator, 5 pole Din
Page 60 A)	HYSTEROSCOPE TELESCOPES STANDARD –	
a.	Operating and Contact-Hysteroscope Forward-Oblique Full HD Telescope 30°, enlarged view, magnification 1x, 60x, diameter 4.0 mm, length 30 cm, autoclavable, fiber optic light transmission incorporated,- 1 no	Operating Hysteroscope Forward-Oblique Full HD Telescope 30°, enlarged view, magnification 1x, 60x, diameter 4.0 mm, length 30 cm, autoclavable, fiber optic light transmission incorporated,- 1 no

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

1. TC compliance sheet is being uploaded on CPP Portal for necessary actions of interested Bidders.
2. Price Bid Format is being uploaded on CPP Portal for necessary actions of interested Bidders.
3. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.