

Response To Pre-Bid Queries			
GEM/2023/B/3447807, Dated: 16.06.2023			
Item Name: Anaesthesia Workstation			
Tender Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
8	The Unit must have emergency oxygen flush . Flow between 25 to 75 liters per minute	The Unit must have emergency oxygen flush. Flow between 30 to 70 liters per minute . <b>Justifications:</b> The asked point is limiting and restricting the competition. Hence, we request you to amend this point for wider participation.	The Unit must have emergency oxygen flush .
B (A)	Ventilation Modes : VCV,PCV,SIMV-VC,SIMV-PC, CPAP, PRVC, SIMV_PRVC,PSV	Ventilation Modes: VCV,PCV,SIMV-VC,SIMVPC, PSV. <b>Justifications:</b> The asked point is limiting and restricting the competition as these modes are available with higher version. Hence, we request you to amend this point for wider participation	No Change

