

**Amendment No. 2****21.01.2023****Sub: Amendment to the Bidding Document****Ref.: Notice Inviting Bid ref. HITES/PCD/MCH-AIIMS/14/22-23 dated 14.12.2022 read with Amendment no 1 dtd. 10.01.2023**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

**SECTION - I****NOTICE INVITING BIDS (NIB)**

<b>Description</b>	<b>Existing</b>	<b>Amended as</b>
Closing date & time for submission of tender fee and <b>EMD in physical form</b>  Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the mentioned date and time	30.01.2023 at 02:00 PM	<b>31.01.2023</b> at 02:00 PM

**Important Note to Prospective Bidders:**

In reference to the extension of bid opening date, participating bidders, who had submitted or are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD as per tender requirement is to be kept in line with the extended bid opening date.

**SECTION – VII****TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 1 (Tender ID: 2022\_HLL\_138217\_1)****Automated Tissue Processor**

<b>Sr. No.</b>	<b>Tender Page &amp; Para</b>	<b>Existing Tender Specification</b>	<b>Amended as</b>
1	Pg 46 Para 26	Open machines will be preferred. However, if machines need specific reagent/s for a validated run of the xylene-free protocol, the cost of the reagents must be separately quoted. The reagent cost would be fixed for the next 5 years.	<b>Deleted</b>
2	NA	<b>Added Para:</b>	<b>BOQ:</b> <b>1. Automated Tissue Processor as specified - 01 No.</b> <b>2. UPS as specified - 01 No.</b>

**Note:**

Based on the above amendment, Technical Compliance Sheet 'TC\_2022\_HLL\_134710\_1\_Ver1.0' has been provided in CPP portal. The same need to be downloaded for submission by participating bidders.

**Item No. 2 (Tender ID: 2022\_HLL\_138217\_2)****PCR Machine**

<b>Sr. No.</b>	<b>Tender Page &amp; Para</b>	<b>Existing Tender Specification</b>	<b>Amended As</b>
1	Pg 47 Para 2	Choice of interchangeable reaction modules should include gradient enabled dual 48/48 well fast, gradient enabled 96-well fast, and gradient enabled 96- deep well and gradient enabled 384- well reaction modules.	Choice of interchangeable reaction modules should include gradient enabled dual 48/48 well <b>x 0.2ml or (3 x 32) x 0.2ml with independent controls</b> fast, gradient <b>or user defined (Veriflex Technology)</b> enabled 96-well fast, and gradient enabled 384- well reaction modules.

<b>Sr. No.</b>	<b>Tender Page &amp; Para</b>	<b>Existing Tender Specification</b>	<b>Amended As</b>
2	Pg 47 Para 3	Fully adjustable heated lids should accommodate a broad range of vessels	Fully adjustable heated lids should accommodate a broad range of vessels/ <b>Block configuration</b>
3	Pg 47 Para 7	Accuracy: $\pm 0.2^{\circ}\text{C}$	Accuracy: <b>Should be in between <math>\pm 0.2^{\circ}</math> - <math>0.3^{\circ}\text{C}</math> or better</b>
4	Pg 47 Para 8	Uniformity: $\pm 0.4^{\circ}\text{C}$ well-to-well within 10 sec	Uniformity: <b>In between <math>\pm 0.4^{\circ}</math> - <math>0.5^{\circ}\text{C}</math> or better</b> well-to-well within <b>20</b> sec
5	NA	<b>Added Para:</b>	<b>BOQ: 1. PCR Machine as specified - 01 No.</b>

**Note:**

Based on the above amendment, Technical Compliance Sheet '**TC\_2022\_HLL\_134710\_2\_Ver1.0**' has been provided in CPP portal. The same need to be downloaded for submission by participating bidders.

**All other contents of the Bidding Document including terms & conditions remain unaltered.**