

Amendment No. 05**Sub: Amendment to the referred tender enquiry**

Ref.: Tender Enquiry Document HITES/PCD/PMSSY-IV/25/Mix/2022-23 dated 12-08-2022 and its Amendment No. 01 dated 18-Aug-2022, Amendment No. 02 dated 19-Sep-2022, Amendment No. 03 dated 12-Oct-2022 & Amendment No. 04 dated 15-Oct-2022.

Following changes are being incorporated in the tender:

SECTION VIITECHNICAL SPECIFICATIONS

For Tender ID 2022_HLL_125318_5 only

Item No 5 Ultrasonic Cutting & Coagulation, Advanced Radio Frequency/ Bipolar Energy and Monopolar Energy System (CTVS)		
Tender spec Para	TENDER SPECIFICATION	READ AS
A.8	Should be able to power energy instruments with microprocessor controlled bipolar electrosurgical radiofrequency technology with a sinusoidal forced impedance output	Should be able to power energy instruments with microprocessor controlled bipolar electrosurgical radiofrequency technology with a sinusoidal I/Quasi sinusoidal forced impedance output
A.12	Should be compatible with integrated RF/bipolar and ultrasonic energy instruments which can simultaneously seal and transect vessels 7mm, large pedicles and vascular bundles with ultrasonic energy with system feedback	Should be compatible with integrated RF/bipolar and ultrasonic energy instruments which can simultaneously seal and transect vessels 7mm, large pedicles and vascular bundles with ultrasonic/ Integrated Ultrasonic & RF energy with system feedback
C.1	Generator cart - 1 no	OEM Generator cart - 1 no.
		<p>Added Para: Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model.</p> <p>In case the supplier has not yet obtained import/manufacturing license from CDSCO for the quoted model, proof of application for CDSCO medical device license to be submitted in the bid document and valid CDSCO license to be produced before award of contract, if applicable.</p>
		<p>Added Para: All Consumables and disposables rates must be quoted in bid and shall remain fixed for next three years from the date</p>

HLL Infra Tech Services Limited

Item No 5 Ultrasonic Cutting & Coagulation, Advanced Radio Frequency/ Bipolar Energy and Monopolar Energy System (CTVS)		
Tender spec Para	TENDER SPECIFICATION	READ AS
		of the installation. The Levied taxes will be as per prevailing rate as specified time to time by government.

Note:

- i. All other contents of the tender enquiry including terms & conditions remain unaltered.
- ii. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.
- iii. EMD should be valid up to 225 days or more from the date of submission mentioned in TED & its subsequent amendments.
- iv. **For Tender IDs Tender IDs 2022_HLL_125318_5 only:**
TC compliance sheet v1.0 and Price bid format v1.0 is being uploaded on CPP Portal for necessary actions of interested Bidders.