

Amendment No. 5

06.04.2023

Sub: Amendment to the Bidding Document

Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/48/22-23 dated 09.01.2023 read with Amendment no 1, 2, 3 & 4 dtd. 07.02.2023, 27.02.2023, 25.03.2023 & 05.04.2023 respectively.

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION – VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 1 (Tender ID: 2023_HLL_140698_1)****IORT Machine (Electron Based)**

Sr. No.	Tender Specification Page & Para	Tender Specification	Amended as
1	Page 47	Page No 47 Standards, Safety Clause number 6 (sub clause I) Equipment standard and safety should comply with the national regulatory AERB guidelines and offered model should have AERB type approval and NOC.	Bidder should provide AERB type approval and NOC before installation of equipment.
2	NA	Added Para:-	Offered model should be compatible with mobile/portable CT scanner.
3	NA	Added Para:-	The IORT vendor should provide IORT management system/ similar to OIS to automatically verify and record treatment results after successful delivery.

Note:

- (i) Based on the above amendment, Technical Compliance Sheet 'TC_2023_HLL_140698_1_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2023_HLL_140698_1_Ver1.1'. The same need to be downloaded for submission by participating bidders.
- (ii) Based on the above amendment, Price Bid Format 'PriceBid_2023_HLL_140698_1_Ver1.0' in CPP portal has been superseded by a new version i.e. 'PriceBid_2023_HLL_140698_1_Ver1.1'. The same need to be downloaded for submission by participating bidders.

Item No. 2 (Tender ID: 2023 HLL 140698 2)**IORT Machine (X-Ray Based)**

Sr. No.	Tender Specification Page & Para	Tender Specification	Amended as
1	3 (i) Pg 50	Spherical Applicator set including spherical applicators from 1 to 5cms with variable diameters for different incision length. The applicators are reusable and sterilizable.	Spherical Applicator set including spherical applicators from 1.5 to 5cms with variable diameters for different incision length. The applicators are reusable and sterilizable.

Note:

- (i) Based on the above amendment, Technical Compliance Sheet 'TC_2023_HLL_140698_2_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2023_HLL_140698_2_Ver1.1'. The same need to be downloaded for submission by participating bidders.
- (ii) Based on the above amendment, Price Bid Format 'PriceBid_2023_HLL_140698_2_Ver1.0' in CPP portal has been superseded by a new version i.e. 'PriceBid_2023_HLL_140698_2_Ver1.1'. The same need to be downloaded for submission by participating bidders.

Item No. 3 (Tender ID: 2023 HLL 140698 3)**Mobile CT Scanner for IORT**

Sr. No.	Tender Specification Page & Para	Tender Specification	Amended as
1	1.11 Pg 53	The system should be compatible with OT table for precise imaging in OR	The system should be compatible with radiolucent OT table top or have an option for integration with the OT table for precise imaging in the OR

Note:

- (i) Based on the above amendment, Technical Compliance Sheet 'TC_2023_HLL_140698_3_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2023_HLL_140698_3_Ver1.1'. The same need to be downloaded for submission by participating bidders.
- (ii) Based on the above amendment, Price Bid Format 'PriceBid_2023_HLL_140698_3_Ver1.0' in CPP portal has been superseded by a new version i.e. 'PriceBid_2023_HLL_140698_3_Ver1.1'. The same need to be downloaded for submission by participating bidders.

SECTION - VIII**QUALIFICATION CRITERIA****Existing Clause:**

2. The Bidder should have supplied and installed at least 1 (one) unit in last Five years from the date of Bid Opening, similar equipment meeting major parameters of technical specification which is functioning satisfactorily.

Amended as:

2. The Bidder **or its manufacturer** should have supplied and installed at least 1 (one) unit in last Five years from the date of Bid Opening, similar equipment meeting major parameters of technical specification which is functioning satisfactorily.

All other contents of the Bidding Document including terms & conditions remain unaltered.