

### **Amendment no. 3**

**Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.**

**Ref.: GeM Bid No GEM/2022/B/2577791 Dated: 26-09-2022 for item 'Defibrillator with disposable pads & pacing facility with feedback (Qty: 09 nos.)'**

The following changes are being incorporated in the above referred GeM Bid Number only.

<b>Sr. No.</b>	<b>Buyer uploaded ATC document Page &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
1	Pg 1 Para 9	It should have ability to measure chest compression rate and depth in real time with visual feedback on screen with rate and depth indicator.	<b>Deleted</b>
2	Pg 1 Para 14 d.	12 lead	<b>5/6 lead</b>
3	Pg 1 Para 15 e)	Reusable CPR feedback sensor/or similar product reused at least on 90 patients – 2 nos	<b>Deleted</b>

#### **Important Note:**

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.
2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.
3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.
4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME
5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.