

Amendment No. 6**14.12.2022****Sub: Amendment to the Global Tender Enquiry Document****Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/46/22-23 dated 03.08.2022 read with its Amendment no. 01, 02, 03, 04 & 05 dtd. 29.08.2022, 15.09.2022, 10.10.2022, 31.10.2022 & 23.11.2022 respectively.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION – I**NOTICE INVITING BIDS (NIB)**

Description	Existing	Amended as
Last date and time of submission of tender:	19.12.2022 at 02:00 PM	26.12.2022 at 02:00 PM
Date and time of tender opening:	20.12.2022 at 02:30 PM	27.12.2022 at 02:30 PM

Important Note to Prospective Bidders:

In reference to the extension of bid opening date, participating bidders, who had submitted or are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.

SECTION – VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 1 (Tender ID: 2022_HLL_124557_1)****CO2 Laser**

Sr. No.	Tender Specification Page & Para	Existing Tender Specification	Amended as
1	Pg 45 Para 16 (A)	It Should Have At Least 50 User Defined Memory settings.	It Should Have At Least 05 User Defined Memory settings.
2	Pg 45 Para 22 (A)	Laser Compatible Endotracheal Tubes Of Different Sizes (5, 5.5 , 6, 6.5,7, 7.5,8, 8.5) 10 each Size	Laser Compatible Endotracheal Tubes Of Different Sizes (5, 5.5, 6, 6.5,7, 7.5,8, 8.5) 10 each. Size can be from other OEM
3	Pg 45 Para 2 (B)	It Should Be Easily Adjustable And Should Have Variable Working Distance From 200 mm To 450 mm.	It Should Be Easily Adjustable And Should Have Variable Working Distance From 225 mm To 500 mm.
4	Pg 46 Para 10(B)	It Should Have A Power Transmission Of Greater Than 90%, With Unlimited Power input.	DELETED
5	Pg 46 Para 1(c)	230 mm Handpiece Unit (Cvd Optical Unit, Ports Holder, M Conical Main Extender, Contamination Collector)	200 mm Handpiece Unit (Cvd Optical Unit, Ports Holder, M Conical Main Extender, Contamination Collector)
6	Pg 46 Para 2 (D)	Straight Rigid Hand Straight Tip 60 mm- 1	Straight Rigid Hand Straight Tip 60-70 mm- 1
7	Pg 46 Para 3 (D)	Straight Rigid Hand Straight Tip 180 mm- 1	Straight Rigid Hand Straight Tip 100-180 mm- 1
8	Pg 46 Para 6 (D)	Bent Rigid Hand Curved Tip 240 mm- 1	Bent Rigid Hand Curved Tip 240-300 mm- 1
9	Pg 46	Price of all required accessories and disposable should be quoted in Price Bid for future use.	DELETED

Important Note to Prospective Bidders:

- (i) Based on the above amendment, Technical Compliance Sheet 'TC_2022_HLL_124557_1_Ver1.0' in CPP portal has been superseded by a new

version i.e. 'TC_2022_HLL_124557_1_Ver1.1'. The same need to be downloaded for submission by participating bidders.

- (ii) Based on the above amendment, Price Bid Format 'PriceBid_2022_HLL_124557_1_Ver1.0' in CPP portal has been superseded by a new version i.e. 'PriceBid_2022_HLL_124557_1_Ver1.1'. The same need to be downloaded for submission by participating bidders.

Item no. 2 (Tender ID: 2022 HLL 124557 2)

Operating Microscope – Advance

Sr. No.	Tender Specification Page & Para	Existing Tender Specification	Amended as
1	Pg 47 Para 9 (Under Microscope Body and Optics)	Should have Laser focusing system for quick focusing	Microscope should have auto focus / Laser focusing system for quick focusing.
3	Pg 47 Para 1(Under ILLUMINATION SYSTEM)	Two coaxial xenon illumination of about 300 W, with back up similar rating xenon with quick-action lamp changer in case of failure of main lamp should be integrated within the microscope stand.	Microscope should have illumination of about 300 W, with back up similar rating xenon with quick-action lamp changer in case of failure of main lamp should be integrated within the microscope stand.
5	Pg 48 Para 7	(Under INTEGRATED DIGITAL VIDEO CAMERA SYSTEM) Camera should be independent of microscope.	Deleted
6	Pg 48 Para 1 (Under ADDITIONAL FEATURES)	Image guidance – Microscope should be fully ready for image guidance system integration.	Deleted
7	Pg 48 Para 2 (Under ADDITIONAL FEATURES)	Microscope should have the facility of image injection for endoscope, CT and MRI images during image guided surgery.	Deleted

Important Note to Prospective Bidders:

Based on the above amendment, Technical Compliance Sheet 'TC_2022_HLL_124557_2_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2022_HLL_124557_2_Ver1.1'. The same need to be downloaded for submission by participating bidders.

Item no. 3 (Tender ID: 2022 HLL 124557 3)**Comprehensive ICG Fluorescence Imaging system**

Sr. No.	Tender Specification Page & Para	Existing Tender Specification	Amended as
1	Pg 49 Para 1 (1.2)	The VIS light source shall be consisted of light emitting diode array.	The VIS Light source shall be consisted of light emitting diode array /LED based
2	Pg 49 Para 1 (1.3)	The NIR light source shall be consisted of NIR laser diode array.	The NIR light source shall be consisted of NIR laser diode array/ LED based NIR light
3	Pg 49 Para 1.12	The device shall have buttons at the front panel for quick operation and setting purpose:	Preferably device shall have buttons at the front panel for quick operation and setting purpose:
4	Pg 49 Para 1.13	The device shall have indicators at the front panel to indicate the following conditions:	Preferably device shall have indicators at the front panel to indicate the following conditions:
5	Pg 50 Para 1.21.2	Repetition rate: 20 pulse/ sec	DELETED
6	Pg 50 Para 1.21.4	Beam divergence: 75 Embedded Laser Source Classification: Class 4	DELETED
7	Pg 50 Para 2.1.6	The camera head shall have an indicator showing laser on/off	DELETED
8	Pg 50 Para 3	Full HD Portable Handheld imager- qty 1 unit	Portable hand held image capture device

Important Note to Prospective Bidders:

- (i) Based on the above amendment, Technical Compliance Sheet 'TC_2022_HLL_124557_3_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2022_HLL_124557_3_Ver1.1'. The same need to be downloaded for submission by participating bidders.
- (ii) Based on the above amendment, Price Bid Format 'PriceBid_2022_HLL_124557_3_Ver1.0' in CPP portal has been superseded by a new version i.e. 'PriceBid_2022_HLL_124557_3_Ver1.1'. The same need to be downloaded for submission by participating bidders.

Item no. 4 (Tender ID: 2022 HLL 124557 4)**ENT Treatment Unit with Camera and Screen**

Sr. No.	Tender Specification Page & Para	Existing Tender Specification	Amended as
1	Pg 53 Para K (1. Under main unit))	Separate stainless steel/Plastic tank to prevent mineral build up and heat up to 38 degree temp. Water storage capacity of 5 litres or more.	Separate stainless steel/Plastic tank to prevent mineral build up and heat up to 37 degree temp. Water storage capacity of 2 litres or more.
2	Pg 53 Para L (1. Under main unit)	Cold water irrigation through existing water connection/ water tank with storage capacity of 5 litre or more	Cold water irrigation through existing water connection/ water tank with storage capacity of 4.5 litres or more
3	Pg 53 Para E (3. Under ENT PATIENT EXAMINATION CHAIR)	Should have complete rotation 360 degree with locking device	Should have complete rotation 340 degree with locking device
4	Pg 54 Para B (5. Under Endoscopy Recording System)	Camera Head& Coupler Should be one Piece	Camera head and coupler should be compatible with ENT endoscopes.
5	NA	Suggested by bidder	Added Para:- 1 drawer equipped with UV sterilizer
6	Page 54, Para 10	During the warranty period and CAMC period, desired Uptime of 95% of 365/366(Leap Year) days (24 Hrs), if downtime more than 5%, the warranty period/CAMC period will be extended by twice the time period. In addition a penalty equal off 0.25% of the total cost of equipment per day will be liveable for the excess downtime period. Complaints should be attended properly maximum within 8 hrs.	DELETED

Sr. No.	Tender Specification Page & Para	Existing Tender Specification	Amended as
7	Page 54, Para 12	The Principle/Supplier should maintain and support the system for upcoming 10 years (warranty + CAMC) if the existing system or any other part are obsolete then principle/supplier upgrade it without any additional cost.	DELETED
8	Page 54, Para 12	Should mention the price for consumable/accessory/spare.	DELETED

Important Note to Prospective Bidders:

Based on the above amendment, Technical Compliance Sheet 'TC_2022_HLL_124557_4_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2022_HLL_124557_4_Ver1.1'. The same need to be downloaded for submission by participating bidders.

Item no. 6 (Tender ID: 2022 HLL 124557 6)

Semi-Rigid Ureterorenoscope with Accessories 6/7.5fr

Sr. No.	Tender Specification Page & Para	Existing Tender Specification	Amended as
1	Pg 56 Para 8.a	Durable detachable quick release bridge for irrigation with two lateral irrigation ports and two working channels	bridge for irrigation with two lateral irrigation ports and Two working channels
2		Must have US FDA and CE certificate	Equipment should be USFDA/European CE/BIS certified.

Important Note to Prospective Bidders:

Based on the above amendment, Technical Compliance Sheet 'TC_2022_HLL_124557_6_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2022_HLL_124557_6_Ver1.1'. The same need to be downloaded for submission by participating bidders.

Item no. 9 (Tender ID: 2022 HLL 124557 9)**Endoscopic Sinus Surgery set**

Sr. No.	Tender Specification Page & Para	Existing Tender Specification	Amended as
1	Pg 61 Para 28	Endoscope sheath-01	DELETED
2	Pg 61 Para 35	Endoscopic Liga clip applicator 100,200-01 each	Endoscopic vessel clip applicator 100,200-01 each
3	Pg 61, para 8	The Instruments is CE/US-FDA approved.	Added Para :- Instrument set should be European CE with 4 digit notified body no/ USFDA/BIS certified
4	NA		Added Para :- 10% variation in sizes is acceptable
5	NA		Added Para:- Bidder should quote for all the instruments of the set. Atleast 75% of the quoted instruments should be from same manufacturer and 25% can be from different manufacturer

Important Note to Prospective Bidders:

- (i) Based on the above amendment, Technical Compliance Sheet 'TC_2022_HLL_124557_9_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC_2022_HLL_124557_9_Ver1.1'. The same need to be downloaded for submission by participating bidders.
- (ii) Based on the above amendment, Price Bid Format 'PriceBid_2022_HLL_124557_9_Ver1.0' in CPP portal has been superseded by a new version i.e. 'PriceBid_2022_HLL_124557_9_Ver1.1'. The same need to be downloaded for submission by participating bidders.

All other contents of the Global Tender Enquiry Document including terms & conditions remain unaltered.