

Response To Pre-Bid Queries (Pre-Bid date: 18.04.2022)				
Bid Number: GEM/2022/B/2079412 Dated: 31-03-2022				
Item Name: Neonatal Ventilator				
Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 3	Basic modes (IMV, SIMV, assist control, PSV, CPAP, NIPPV) as well as hybrid modes such as volume targeted modes (VG/TTV/PRVC) should be available	Basic modes (IMV, SIMV, assist control, PSV, CPAP, NIPPV) as well as hybrid modes such as volume targeted modes (VG/TTV/PRVC). Please add: SNIPPV and NHFOV	No change
2	Pg 1 Para 4	High frequency oscillation mode with volume targeting or volume guarantee option also should be available	We request you to kindly delete this point Remarks/Justification: Kindly note that mentioned features are available with one or two companies worldwide, which is limiting competition. So, in order to have better competition please delete this point. We request to amend as "High frequency oscillation mode in PRVC Mode "	No change
3	Pg 1 Para 7 Capable of delivering:	PIP: 0-60 cm water	PIP: 5-60 cm water Justification: For wider participation	No change
4	Pg 1 Para 7 Capable of delivering:	MAP: up to 45 cm H ₂ O	MAP: up to 30 cm H ₂ O or more Justification: For wider participation	No change
5	Pg 1 Para 8	Inbuilt digital display – TFT/LCD of screen size at least 10 inches:	We request to amend as TFT 15" Screen	No change
6	Pg 1 Para 12	Should be compatible with inhaled nitric oxide (iNO) system of any brand	Kindly delete this point. Remarks/Justification: The requested specs are available with few companies only worldwide, which is limiting the competition. So, in order to have better and wider competition please amend this point. Should be compatible with inhaled nitric oxide (iNO) system of any brand and certificate to be provided from INO manufacturer.	No change

Response To Pre-Bid Queries (Pre-Bid date: 18.04.2022)				
Bid Number: GEM/2022/B/2079412 Dated: 31-03-2022				
Item Name: Neonatal Ventilator				
Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
7	Pg 1 Para 13	Inbuilt nebulizer	Inbuilt Nebulizer/Integrated with wiremess technology.	No change
8	Pg 1 Para 15	Battery backup for at least 30 minutes	We request to amend as Battery backup 180 minutes.	No change
9	Pg 2 Supplies	F & P MR 850 (including heater wire adaptor): 2 with each unit	Humidifier (including heater wire adaptor): 2 with each unit Justification: Kindly remove company specific terms for wider participation	No change
10	Pg 2 Supplies	Pressure/flow sensor: 20 with each unit	Please clarify - Reusable/Disposable Reusable : 2 nos. Disposable : 40 nos.	Amended as: Reusable Pressure/flow sensor: 10 with each unit Warranty/CMC should include replacement of flow sensors as and when required throughout the warranty/CMC period
			Reuseable flow sensor: 02 with each unit Justification: One reusable flow sensor can be used around 70 - 80 times. Kindly allow reuseabke flow sensor.	
11	Pg 2 Supplies	Flow sensor cable: 3 with each unit	Flow sensor cable: 3 with each unit. It should be cover under warranty.	Amended as: Flow sensor cable: 3 with each unit Warranty/CMC should include replacement of flow sensor cable as and when required throughout the warranty/CMC period
			Should be removed Justification: Kindly remove company specific terms for wider participation	
12	Pg 2 Supplies	Spare oxygen cell: 5 (total)	Spare oxygen cell : 5 (total). It should be cover under warranty.	Amended as: Spare oxygen cell: 5 (total) Warranty/CMC should include replacement of spare oxygen cells as and when required throughout the warranty/CMC period
			Spare oxygen cell: should be supplied as and when it expires during wararnty and CMC period Justification: Oxgen Sensor has limited shelf life . Better , it should be covered under warranty and CMC period .	
13	NA	Suggested by Bidder	Should have HFO and HFO+IMV mode	The specifications include the number of sigh breaths required. It implies that HFO+IMV is required.