

27.12.2021

Amendment No. 7

Sub: Extension of due dates to the referred tender enquiry

Ref.: HITES/PCD/AIIMS-IV/45/Core-Lab/21-22 dated 22.10.2021 read with its Amendment no 1, 2, 3, 4, 5 and 6 dated 28.10.2021, 17.11.2021, 24.11.2021, 25.11.2021, 03.12.2021 and 20.12.2021

The following changes are being incorporated in the above referred Tender Enquiry Document.

**SECTION VII
Technical Specification**

**Tender Ref No : HITES/PCD/AIIMS-IV/45/Core-Lab/21-22 dated 22.10.2021
(Tender ID : 2021_HLL_92294_1)**

Para No	Existing Specification	Amended as :
2.1	Clinical Chemistry System	2.1 Clinical Chemistry System
1	Two fully automated, random access, floor model biochemistry analyzers integrated to track based lab automation and each analyser should have a minimum throughput of 1200 tests/hour with ISE.	Two fully automated, random access, floor model biochemistry analyzers integrated to track based lab automation and each analyser should have a minimum throughput of 800 tests/hour .

**All other contents of the Tender enquiry including terms & conditions remain unaltered.
Note:**

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**