Response To Pre-Bid Queries (Pre-Bid date: 21.12.2022)				
GeM Bid Number: GEM/2022/B/2866765 Dated: 16-12-2022				
Item Name: Adult Multipara Monitor				
Sr. No.	**Buyer Specification Document para and pg**	**Tender Specification**	**Representation received from Bidder**	**COMMITTEE RECOMMENDATION**
1	Para 2.2, page 1	Capability of storage of patient data and printing of patient reports	Request to amend as-please insert USB port for patient data download option. Reason: It becomes convenient to use the monitor during the data transfer process and it would be low power consumption.	Facility to patient data download.
2	Para 3.2, page 1	Should have facility to monitor and display - ECG, NIBP,SpO2(Nellcor/Masimo), Temperature, Respiration and upgradable to ETCO2(Sidestream or Microstream) & IBP.	Please inset original Nellcor or Masimo accessories to be provided and compatible accessories will not be accepted. Reason: As we know Nellcor/ Masimo both are golden standard of pulse oximetery techhnology and it provides fast, accurate and consistent pulse oximetery performance. it will minimise motion artifacts as well.	Should have facility to monitor and display - ECG, NIBP,SpO2 **(Original Nellcor or Masimo technology)**, Temperature, Respiration and upgradable to ETCO2(Sidestream or Microstream) & IBP.
3	Para 3.9, page 1	Should be capable of measuring oxygen Saturation even in case of motion artifact.	Request you to remove this specification for wider participation.	No change Considered.
4	Para 3.13, Page 1	Should have inbuilt 2 Ch thermal recorder with selectable recording speed of 25 & 50 mm /sec.	Request to amend as -should have recording speed of 25 mm/sec Reason: The standard output speed is 25mm per second. If a different paper speed is used, standard rate calculations will have to be modified appropriately. We request you to amend the specification for Wider Participation.	Should have inbuilt 2 Ch thermal recorder with selectable recording speed of atleast 25mm /sec.
5	Para 5.1, page 1	Environmental factors: The unit shall be capable of operating continuously in ambient temperature of 0 -40° C and relative humidity of 15-90%	The unit shall be capable of operating continuously in ambient temperature of 5 -40° C and relative humidity of 15-85% Reason: Kindly allow for wider participation	Environmental factors: The unit shall be capable of operating continuously in ambient temperature of 5-40° C and relative humidity of 15-85%.
6	NA	Suggested by Bidder	**Request to add:** monitor should have trend storage of 168 hours Reason: This will help in storing more information .	Not Considered

Response To Pre-Bid Queries (Pre-Bid date: 21.12.2022)				
GeM Bid Number: GEM/2022/B/2866765 Dated: 16-12-2022				
Item Name: Adult Multipara Monitor				
Sr. No.	**Buyer Specification Document para and pg**	**Tender Specification**	**Representation received from Bidder**	**COMMITTEE RECOMMENDATION**
7	NA	Suggested by Bidder	**Request to add:** monitor should have trend storage on PEN drive / SD Card. And the data should be easily accessible in computer without help of any special software Reason: for MRD records it would be helpful	Not Considered
8	NA	Suggested by Bidder	**Request to add:** Should have USB connectivity directly to laser printer Reason: so that if required a plain paper printout can be taken.	Not Considered
9	NA	Suggested by Bidder	**Request to add:** Should have 12 lead ECG monitoring facility by 10 lead ECG cable. Reason: 12 lead ECG monitoring will help in proper analysis of ischemia related disease	Not Considered
10	NA	Suggested by Bidder	**Request to add:** Should have minimum 12 waveform on display Reason: Helps in veiwing of all 12 lead ECG in one single screen	Not Considered
11	NA	Suggested by Bidder	**Request to add:** Should have softwares lik ST analaysis, ST mapping, QT analysis, HRV analysis and should measure PVI & PI Reason: for proper diagonosis these features are required in a multiparamonitor	Not Considered
12	Para 8.4, pg 2	List of important spare parts and accessories with their part number and costing and to be blocked for 5 years.	as per prebid discussion	Deleted
13	Para 7.1, pg 1	Should be USFDA or European CE with 4 digit notified body no. or BIS Certified.	as per prebid discussion	Monitor and should be European CE with four digit notified body number or US FDA or BIS approved for the quoted model and certificate to be submitted. OR Monitor should meet IEC 60601-1, IEC 60601-1-2 & IEC 80601-2-49 standards and submit valid test report from any NABL accredited labs or from the labs in their country of origin (incase of foreign manufacturers) for the quoted model)

Response To Pre-Bid Queries (Pre-Bid date: 21.12.2022)				
GeM Bid Number: GEM/2022/B/2866765 Dated: 16-12-2022				
Item Name: Adult Multipara Monitor				
Sr. No.	**Buyer Specification Document para and pg**	**Tender Specification**	**Representation received from Bidder**	**COMMITTEE RECOMMENDATION**
14	Para 8.3, pg 2	Compliance Report to be submitted in a tabulated and point wise manner clearly mentioning the Page /para number of original catalogue / data sheet. Any point, if not substantiated with Authenticated catalogue/manual, will not be considered.	as per prebid discussion	Deleted
15	Para 8.4, pg 2	List of important spare parts and accessories with their part number and costing and to be blocked for 5 years.	as per prebid discussion	Deleted
16	Page 8.5, pg 2	Log book with instruction for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.	as per prebid discussion	Deleted
17	BOQ, PG 2	Added Para	as per prebid discussion	Extra large NIBP cuff - 1 No