10-03-2023

Amendment no. 6

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/2702810 Dated: 03-11-2022 for item 'Digital Flat Panel Fluoro cum Radiography (Qty: 1 no.)

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	Buyer Specification document Para and Page	EXISTING TENDER SPECIFICATION	READ AS
1	Para G (Control console), point 4, pg 1	Added Para : The control software must have latest security windows 10 platform only.	The control software must have latest security windows 10/equivalent compatible platform.
4	Para I (Accessories), pt 5, pg 2	Radiation protection flaps	lead barrier with lead glass for Radiation protection.
5	Para (Site Modification), point 3, pg 2	The cost of site modification for the area of 1000sq.ft and Air-conditioning of Tonnage 20 TR (including standby unit/s) will be considered for Ranking / Evaluation purpose. However, the payment will be paid on pro-rata based on the actual site modification area	The cost of site modification for the area of **350 sq.ft** and Air-conditioning of Tonnage **6TR** (including standby unit/s) will be considered for Ranking / Evaluation purpose. However, the payment will be paid on pro-rata based on the actual site modification area
6	Paar BOQ(site modification work Work), PG 3, POINT 1	Civil works 1000 Sq.ft	**Civil works 350 Sq.ft**
7	Paar BOQ(site modification work Work), PG 3, POINT 2	Electrical work 1000 Sq. ft	**Electrical work 350 Sq. ft**
8	Paar BOQ(site modification work Work), PG 3, POINT 3	Public health (plumbing and sanitary fittings). 1000 Sq.ft.	**Public health (plumbing and sanitary fittings): 350 Sq.ft.**
9	Paar BOQ(site modification work Work), PG 3, POINT 4	Air Conditioning 12 TR	**Air Conditioning: 06 TR**
10	Paar BOQ(site modification work Work), PG 3, POINT 5	Interior Furnishing & Furniture 1000 Sq.ft.	**Interior Furnishing & Furniture 350 Sq.ft.**
11	Paar BOQ(Miscellaneous), PG 3, POINT 5	Toilet (Optional) - 1	Deleted
12	Para BOQ(Miscellaneous), PG 3	All the Optional items offered should be quoted separately and the price should be valid for warranty period	Deleted

13	NA	**Added Para:**	Indicative Layout drawing attached.

Important Note:

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.

2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.

3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.

4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME

5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.