10-03-2023

Amendment no. 6

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/2701599 Dated: 03-11-2022 for item 'C-ARM Digital Flat Panel Fluoroscopy (Under Couch) (Qty: 3 nos)

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	Buyer Specification document Para and Page	EXISTING TENDER SPECIFICATION	READ AS
1	Para A (GANTRY / C-ARM), point 6, pg 1	Depth of immersion: 72 cm or more	Depth of immersion: 65 cm or more.
2	Para D (FLAT PANEL DETECTOR SYSTEM), point 1, pg 2	Flat detector of CsI (Cesium Iodide) with amorphous Silicon doping	Flat detector of CSI with amorphous silicon doping OR CMOS (CMOS Detectors is the most latest technology in detectors. CMOS detectors have best resolution, pixel size, DQE) .
3	Para G (Terms and condition), point 2, pg 3	US-FDA and CE (European) and AERB-type approved for the main fluoroscopy system	BIS/ USFDA/CE (European) with 4 digit and AERB-type approved for the quoted C-arm system. **Added Para:** Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model. In case the vendor has not yet obtained import/manufacturing license from CDSCO for the quoted model, proof of application for CDSCO medical device license to be submitted in the bid document and valid CDSCO license to be produced at the time of supply/ NOA for the quoted model
4	Para G, point 1, page 3	Please attach the original manufacture's product catalogue and datasheets. Computer printed sheets are not acceptable	Please attach the original manufacture's product catalogue and datasheets.

5	Para 9, page 3	Revolving type of executive chairs with castor- Qty-04 Nos, Site fabricated workstation tables with waterproof plyboard - Qty-01Nos ,Site fabricated cupboard with waterproof plyboard (Size- 3'3'' X 7'10'')-Qty-02 Nos, Instrument trolleys- 1 No, Patient footstep – 1 no, Drug trolley – 1 no, Patient trolley with rubber foam mattress to be kept in the patient preparation room- 1 No, Dustbins (plastic with lid) to be provided- 3 No, CCTV camera with 2GB DVR, LCD monitor for facility	Revolving type of executive chairs with castor- Qty-04 Nos, **Suitable** workstation tables with waterproof plyboard - Qty-01Nos , Wall mounted cupboard with waterproof plyboard -Qty-02 Nos, Instrument trolleys- 1 No, Single Patient footstep – 1 no, **Double foot step-1 No,** Drug trolley – 1 no, Patient trolley with rubber foam mattress to be kept in the patient preparation room- 1 No, Dustbins (plastic with lid) to be provided- 3 No, CCTV camera with **1TB D**VR, LCD monitor for facility.		
6	Para 10, pg 1	Equipped with two high-resolution 18" HD LED/LCD/TFT monitors with Diagnostic Image quality with image matrix of 1024 x 1024 or more	Equipped with two Full HD 21" LCD/LED monitors with Diagnostic Image quality.		
7	NA	Added Para	The cost of Site Modification Work for the area of 370 sq.ft and Air-conditioning of Tonnage 5TR (including standby unit) will be considered for Ranking /Evaluation purpose		
8	Para F(Power Specifications), pg 2	Suitable Ac to be supplied along with system - 6 Tr.	Suitable Ac to be supplied along with system - 5 Tr.		
9	NA	NA	**Added Para: BOQ**		
		1	C-Arm Digital Flat Panel Fluoroscopy	1	No
		2	Online UPS		
		3	18" HD Monitors	2	Nos
			Accessories		
		1	Zero Lead Aprons	5	Nos
		2	Thyroid Shield	5	Nos
		3	Sterile covers for C-arm	10	Nos
		4	Gonadal shields	5	Nos
		5	Lead goggles	5	Nos
		6	Table as per specification	1	No
			site modification work Work (370 sq ft) as per specification	370	Sq. Ft.
		1	Air Conditioning	5	TR
			Furniture:		
		1	Executive chairs	4	Nos
		2	workstation tables	1	No
		3	cupboard	2	Nos
		4	Instrument Trolley	1	No
		5	Footstep	1	No

		6	Drug Trolley	1	No
		7	Patient Trolley	1	No
		8	Dustbin	3	Nos
		9	CCTV Camera	1	No

Important Note:

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.

2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.

3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.

4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME

5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.