Amendment no. 6

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/2702566 Dated: 03-11-2022 for item 'Mobile X ray – Flat Panel DR (Qty: 3 nos)

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	Buyer Specification document Para and Page	EXISTING TENDER SPECIFICATION	READ AS
1	Para and pg 1	State-of-the-art, battery driven, compact, easily transportable, USFDA approved digital mobile radiographic units with telescopic or articulated arm and flat panel detector are required. It should be suitable for bedside radiography in wards, intensive care units and operation theatres.	State-of-the-art, battery driven, compact, easily transportable, BIS/USFDA/ European CE approved digital mobile radiographic units with telescopic or articulated arm and flat panel detector are required. It should be suitable for bedside radiography in wards, intensive care units and operation theatres.
2	Para 6(Other Features), Point f, pg 1	Entire unit should be USA FDA approved.	Deleted
3	NA	**Added Para**	3 Light weight aprons should be provided.
4	NA	**Added Para**	3 Thyroid shields to be provided.
5	NA	**Added Para**	3 Gonad shield to be provided.
6	NA	**Added Para**	3 google/eye shield
7	Para , pg 2	State-of-the-art, battery driven, compact, easily transportable, USFDA approved digital mobile radiographic units with telescopic or articulated arm and flat panel detector are required. It should be suitable for bedside radiography in wards, intensive care units and operation theatres	Deleted

8	Para 9, pt c, pg 2	Warranty should be followed by comprehensive annual maintenance contract (CMC) for next five years, covering all items as that of warranty. The Cost of CMC with year wise breakup from 6th to 10th years should be mentioned in the price bid.	Deleted
9	Para 9, pt d, pg 2	95% uptime guarantee should be given. In case downtime exceeds 5%, penalty in the form of extended warranty, double the number of days for which the equipment goes out of service will be applied.	Deleted
10	Para 9, pt e, pg 2	There should be at least three installations of quoted unit in globally. Satisfactory performance report from the existing users should be provided	Deleted
11	Para 6, pt f, pg 1	Entire unit should be USA FDA approved	Deleted
12	Para 12, pg 2	**Added Para**	**Approvals** Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model.In case the vendor has not yet obtained import/manufacturing license from CDSCO for the quoted model, proof of application for CDSCO medical device license to be submitted in the bid document and valid CDSCO license to be produced at the time of supply/ NOA for the quoted model
13	Para 12, pg 2	**Added Para**	The system should be AERB type approved and the copy of E-LORA Listing should be submitted along with bid. If the quoted model has not been yet installed in India, vendor should submit NOC from AERB.Regular QA according to AERB norms will

			be responsibility of bidder during warranty and CMC period.		
14	Para 12, pg 2	**Added Para**	Regular QA according to AERB norms will be responsibility of bidder during warranty and CMC period.		
15	NA	**Added Para**	**BOQ**	**Qty**	**UOM**
		1	Mobile X-ray Flat panel DR ,as per the Tender Specification	1	No
		2	DICOM printer	1	No
		3	Light weight aprons	3	Nos
		4	Thyroid shields	3	Nos
		5	Gonad shield	3	Nos
		6	google/eye shield	3	Nos

Important Note:

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.

2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.

3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.

4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME

5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.