

**Amendment No. 05**

**Sub: Amendment to the referred tender enquiry**

**Ref.: Tender Enquiry HITES/PCD/DHSLD/01/22-23 dated 15-07-2022 (Tender ID: 2022\_HLL\_122468\_1) read with Amendment No. 01 dated 24-08-2022, Amendment No. 02 dated 12.09.2022, Amendment No. 03 dated 23.09.2022 and Amendment No. 04 dated 10.10.2022 respectively**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**Section – VII  
Technical Specifications**

**MRI 1.5 T (2022\_HLL\_122468\_1)**

<b>Tender Page &amp; Para</b>	<b>EXISTING TENDER SPECIFICATION</b>	<b>AMENDED AS</b>
Page 47, Para 4.b	4. Gradient System: b) The gradient should be actively shielded with each axis having independently a <b>slew rate of at least 100 mT/m/ms</b> and a peak amplitude of at least <b>30mT/m</b> at max field of view offered by the scanner	4. Gradient System: b) The gradient should be actively shielded with each axis having independently a <b>slew rate of at least 120 T/m/s</b> and a peak amplitude of <b>33 mT/m</b> at max field of view offered by the scanner.
Page 49, Para 9.f	Coil System: f) General purpose flexible coils - Small & Large with minimum 16 channel each should be quoted. Coil positioning accessories has to be offered	Coil System: f) General purpose flexible coils - Small & Large with minimum 16 channel each should be quoted. Coil positioning accessories has to be offered <b>Added Para: The coils with the highest number of channels available with the vendor for the MRI platform should be quoted.</b>
Page 50, Para 11.b	11. Workstation: b) It should have at least 21-inch LCD TFT 2MP medical grade color monitor, with hard disk of at least 1 TB and 32 GB RAM capacity or more, with self-playing DVD/CD archiving facility. Additional storage of <b>10B</b> has to be provided attached to the workstation	11. Workstation: b) It should have at least 21-inch LCD TFT 2MP medical grade color monitor, with hard disk of at least 1 TB and 32 GB RAM capacity or more, with self-playing DVD/CD archiving facility. Additional storage of <b>10TB</b> has to be provided attached to the workstation
Page 50, Para 15.a	SUITABLE RF ENCLOSURE: a) RF Cabin: The system should be supplied with the <b>imported</b> RF cabin with RF room shielding, RF Door, RF window, and interiors for the same should be carried out suitably	SUITABLE RF ENCLOSURE: a) RF Cabin: The system should be supplied with the RF cabin with RF room shielding, RF Door, RF window, and interiors for the same should be carried out suitably

Tender Page & Para	EXISTING TENDER SPECIFICATION	AMENDED AS
Page 51, Para 19		<b>Added Para:</b> <b>19. Turnkey</b> <b>n. The cost of Site modification work for the area of 1000 sq.ft will be considered for Ranking / Evaluation purpose.</b> <b>o. The actual area of Site modification work done will be considered for payment based on the turnkey rate and site measurements.</b>

All other contents of the tender enquiry including terms & conditions remain unaltered.

**Note:**

- i. In reference to the extension of bid opening date, the participating bidders who had submitted or are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.
- ii. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.
- iii. Technical Compliance Sheet (v1.0) and Price Bid Format (v1.0) is uploaded on CPP Portal along with this amendment for necessary action of interested bidders. This will supersede any version uploaded earlier.