

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575333 Dated: 24-09-2022				
Item Name: Bubble CPAP Machine				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 2	Light weight, portable and sturdy.	Light weight, portable and study with inbuilt battery back and colour display touch screen.	No Change
			Light weight, portable and study with inbuilt battery back and colour display touch screen.	
2	Pg 1 Para 7	Pressure: 3 to 12 cm H2O	2-10 cm H2O (10cm is maximum setting of PEEP in all Bubble CPAP and maximum recommended for Neonatal in Preterm Patients	Amended as: Pressure: 2 to 10 cm H2O or more
			Kindly make it Pressure : 3 to 10 CmH2o Reason for Change: The standard bubble bottle bar has pressure grid 3 cm - 10 cm	
			2-10 cm H2O (10cm is maximum setting of PEEP in all Bubble CPAP and maximum recommended for Neonatal in Preterm Patients	
3	Pg 1 Para 13	FiO2 concentration: 21-100% with accuracy of 5%	Yes (Accuracy should be at least +- 3%	No Change
			Yes (Accuracy should be at least +- 3%	
4	Pg 1 Para 14	Should be BIS or US FDA or European CE with 4 digit notified body no. approved product	Should be made in India as per Govt of India Standards	No Change
			Should be BIS or US FDA or European CE with 4 digit notified body no. Or CDSCO (MDR 2017) / IEC 606001 / ISO-13485 approved product Reason for Change: Many of Indian manufacturers do not have USFDA & CE certifications. So other certifications should be considered for MII mandatory tenders as allowed under clause 7(d) of MDR2017.	
			Should be made in India as per Govt of India Standards	

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575333 Dated: 24-09-2022				
Item Name: Bubble CPAP Machine				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
5	Pg 1 Supplied for each unit Para 1	Humidifier (F&P 850): 1	<p>There are many humidifiers available in market, MR850 is dedicated from F&P, so no other can quote as they offer high prices to competition, due to which there are limited chances for winning bid for other clients who will quote tender. You are kindly requested to put Servo control humidifier word instead of asking specific company brand humidifier.</p> <p>This is a company specific. This should be changed to Servo Controlled Humidifier with USFDA / European CE (with Four digit Notified Number) certification</p> <p>Heated Humidifier preferably MR 850 or equivalent. Reason for Change: MR-850 is proprietary model of M/s Fisher & Paykel. The equivalent model with similar specs should be considered. Since proprietary models cannot be asked for in tenders it becomes mandatory to write as "Heated Humidifier preferably MR-850 or equivalent" so that tender specifications are not restrictive</p> <p>There are many humidifiers available in market, MR850 is dedicated from F&P, so no other can quote as they offer high prices to competition, due to which there are limited chances for winning bid for other clients who will quote tender. You are kindly requested to put Servo control humidifier word instead of asking specific company brand humidifier.</p>	No Change
6	Pg 1 Supplied for each unit Para 3	Spare humidification chamber of the F&P MR 850 humidifier: 4	<p>Kindly remove specific company brand name from bid</p> <p>This is a company specific. This should be changed to Spare humidification chamber for Servo Controlled Humidifier</p> <p>Spare humidification chambers for the humidifier: 4 Reason for Change: Same as above</p> <p>Kindly remove specific company brand name from bid</p>	No Change

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575333 Dated: 24-09-2022				
Item Name: Bubble CPAP Machine				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
7	NA	Suggested by Bidder	Device should be advance Bubble CPAP with option of HFNC, Nasal CPAP and Auto T- Piece with integrated Battery back, integrated Blender.	No Change
			Device should be advance Bubble CPAP with option of HFNC, Nasal CPAP and Auto T- Piece with integrated Battery back, integrated Blender.	
8	NA	Suggested by Bidder	Device should have integrated Air turbine for use during transportation of patients.	No Change
			Device should have integrated Air turbine for use during transportation of patients.	
9	NA	Suggested by Bidder	Device should have display touch screen minimum 4"	No Change
			Device should have display touch screen minimum 4"	