

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2574489 Dated: 24-09-2022				
Item Name: Therapeutic Hypothermia (Whole body cooling device)				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 1	Micro-processor based servo-controlled whole body cooling-warming system for neonates	It may be amended as " microprocessor- controlled cooling Warming System for neonates"	No Change
2	Pg 1 Para 4	Uses high quality single use or reusable blankets of size at least 12"x15"; if reusable, the blanket should be cleanable after each use	It may be amended as " should have facility of gradient / manual adjustment of temperature."	No Change
3	Pg 1 Para 6	Ability to re-warm body to normal temperature gradually (at 0.5oC per hour)	It may be amended as ability to rewarm body to normal temperature "	No Change
4	Pg 1 Para 12	Should be BIS or US FDA or European CE with 4 digit notified body no. approved product	<p>It should be amend with <b>ISO 13485/BIS or USFDA or CE Compliance/European CE</b> approved product</p> <p><b>Note:</b> We have huge installation based at Govt sectors, Military hospital's as well as private hospitals.</p> <p><b>Recently we got order from WBMSCL, Kolkata for IPGME&amp;R and SSKM Hospital. FYR, attached PO copy And SGPGIMS, Lucknow, U.P</b></p> <p>Remarks/ Request/Suggestion: After amendment more and more company can participate in this tender.</p> <p>When the Government of India is promoting  <b>"AATMANIRBHAR BHARAT" &amp; "MAKE IN INDIA"</b>  <b>Public Procurement (preference to Make In India) order 2017</b>  We are looking your support to give chance to Indian Manufacturers.</p>	No Change