

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575366 Dated: 24-09-2022				
Item Name: Neonatal Incubator				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 3 a.	Front and head access door	Front and side access door Justification: For wider participation	No Change
2	Pg 1 Para 20	It should be BIS or US FDA or European CE approved product OR Should meet IEC 60601-1 , IEC 60601-1-2 and IEC 60601-2-20 standards and should submit valid test report from any NABL accredited lab or from the labs in their country of origin (in case of foreign manufacturers) for the quoted model	The equipment should US FDA and CE certified. The supporting document in this regard should be submitted along with the bid. Justification: This is a safety standard feature for Equipment as well as patient	Amended as: It should be BIS or US FDA or European CE approved product
3	NA	Suggested by Bidder	The Unit should be supplied with superior quality pressure diffusing Mattress maintains skin integrity. Justification: It has a pick pressure of 15 mmhg compared to traditional/Gel mattress over 30 mmhg. Note: Tissue nutrition gets effected when critical capillary pressure is >20A-32mmhg	No Change
4	NA	Suggested by Bidder	Should have Integrated servo humidifier. Justification: It is clinically tested against infection for long usages and delivers humidity in sterile form.	No Change