

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)					
Bid Number: GEM/2022/B/2575490 Dated: 24-09-2022					
Item Name: Portable ECMO					
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para A.1 Point 2	Should work within a low prime volume ( 40-80 ml )	M/s B L Lifesciences Pvt. Ltd.	17 ml prime	<b>Amended as:</b> Should work within a low prime volume ( <b>less than 20 mL</b> )
2	Pg 1 Para A.1 Point 4	Flow range 0.1 – 9 litres /min	M/s B L Lifesciences Pvt. Ltd.	0-8 liters	<b>Amended as:</b> Flow range <b>0 - 8 litres /min</b>
3	Pg 1 Para A.1 Point 6	Low profile fins for efficient blood flow of up to 9 L/min	M/s B L Lifesciences Pvt. Ltd.	Upto 8 liters. Magnetic levitation	<b>Amended as:</b> Low profile fins for efficient blood flow of up to <b>8 L/min</b>
4	Pg 1 Para A.2 Point 6	Rotation range 0-3600 rpm	M/s B L Lifesciences Pvt. Ltd.	0-10000	<b>Amended as:</b> Rotation range <b>0-10000 rpm</b>
5	Pg 1 Para A.3 Point 2	Should work within a low prime volume ( 40-80 ml )	M/s B L Lifesciences Pvt. Ltd.	16 ml.	<b>Deleted</b>
6	Pg 1 Para A.3 Point 5	Rotation range 0-3600 rpm	M/s B L Lifesciences Pvt. Ltd.	0-10000	<b>Deleted</b>
7	Pg 1 Para A.4 Point 2	Should generate 0 to 50000 RPM	M/s B L Lifesciences Pvt. Ltd.	0-10000 rpm	<b>Amended as:</b> Should generate 0 to <b>10000 RPM</b>
8	Pg 1 Para A.4 Point 3	Should be able to flow 0.1 to 9.99 Litres per minute.	M/s B L Lifesciences Pvt. Ltd.	0.1-8 liters	<b>Amended as:</b> Should be able to flow <b>0.1 - 8 Litres</b> per minute.
9	Pg 1 Para A.4 Point 4	Should have compact, lightweight console (approx 10 kg weight)	M/s B L Lifesciences Pvt. Ltd.	Approx:12 kgs	<b>Amended as:</b> Should have compact, lightweight console (approx <b>12 kgs</b> weight)
10	Pg 1 Para A.4 Point 5	Should have LCD display screen (5 .5" -7 ")	M/s B L Lifesciences Pvt. Ltd.	10.7 inch TFT touch screen	<b>Amended as:</b> Should have <b>10.7 inch TFT touch screen</b>
11	Pg 2 Para v.	01 no Venous Blood Oxygen Saturation (SvO2) (measuring range – 40 to 100%)	M/s B L Lifesciences Pvt. Ltd.	To be deleted	No Change
12	Pg 2 Para vi.	01 no Hemoglobin (Measuring Range 05 to 15 gm/d)	M/s B L Lifesciences Pvt. Ltd.	To be deleted	No Change

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)					
Bid Number: GEM/2022/B/2575490 Dated: 24-09-2022					
Item Name: Portable ECMO					
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
13	Pg 2 Para vii.	01 no Hematocrit (Measuring range 15 to 50%)	M/s B L Lifesciences Pvt. Ltd.	To be deleted	No Change
14	Pg 3	Cytokine Adsorber	M/s B L Lifesciences Pvt. Ltd.	Not an ECMO kit item	<b>Deleted</b>
15	Pg 3 Cytokine Adsorber Para 1	High flow, low resistance cartridge capable of treating whole blood or plasma with flow rates of 100-400 mL/min using a dialysis or CRRT machine and up to 700 mL/min using a heart-lung machine without hemolysis helpful in countering ongoing cytokine production.	M/s B L Lifesciences Pvt. Ltd.	Different consumable , optional for the end user .	<b>Deleted</b>