

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575280 Dated: 24-09-2022				
Item Name: Radiant Warmer-cum-Incubator (Convertible Unit)				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 2 b.	Humidity: range 5 to 85% relative humidity; should not form any condensation on the walls; accuracy of 10%; reservoir capacity of at least 750-1000 mL	Humidity: range 30 to 95% relative humidity; should not form any condensation on the walls; accuracy of 10%; reservoir capacity of at least 750-1000 mL Justification: For wider Participation.	Amended as: Humidity: range 30 to 95% relative humidity; should not form any condensation on the walls; accuracy of 10%; reservoir capacity of at least 750-1000 mL
2	Pg 1 Para 7	Color display screen – at least 8" size – for displaying key information	Central Color display screen – at least 10" size – for displaying key information Justification: Accessible from both side and easy to visible from distance in NICU.	No Change
3	Pg 1 Para 12	It should be BIS or European CE (with 4 digit notified body no.) or US FDA approved product OR Should meet IEC 60601-1, IEC 60601-1-2 and IEC 60601-2-21 standards and should submit valid test report from any NABL accredited lab or from the labs in their country of origin (in case of foreign manufacturers) for the quoted model	The equipment should US FDA and CE certified. The supporting document in this regard should be submitted along with the bid. Justification: This is a safety standard feature for Equipment as well as patient.	No Change
4	NA	Suggested by Bidder	The Unit should be supplied with superior quality pressure diffusing Mattress maintains skin integrity. Justification: It has a pick pressure of 15 mmhg compared to traditional/Gel mattress over 30 mmhg. Note: Tissue nutrition gets effected when critical capillary pressure is >20A-32mmhg	No Change