

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575469 Dated: 24-09-2022				
Item Name: Heated, Humidified, High Flow Nasal Cannula (HHHFNC)				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 1	Device should have integrated flow generator to deliver flows from 1 to 20 litres.	Please remove Integrated flow generator Device should have flows from 1 to 20 litres. Reason for Change: The point number 1 and point no 7 of specification is contradictory, in point no 9 it is specifically asked for the unit to work on central pipe line. Clarification may be provided.	Amended as: Device should deliver flows from 1 to 20 litres.
2	Pg 1 Para 9	Compressor NOT required– should operate on central compressed air line	Contradictory to first point. The device has built in flow generator hence does not require central compress line. General request to change: Please delete this point	Deleted
3	Pg 1 Para 11	Should be BIS or US FDA or European CE approved product	Should be CE or BIS or US FDA Certified. Should be BIS or US FDA or European CE or CDSCO (MDR 2017), approved product Reason for Change: For make in India products the certification should be considered as per MDR 2017 guideline	No Change
4	Pg 1 Para 12	Certificate of good performance of machine/maintenance services for at least one year from TWO reputed government/private hospital with NICU/PICU bed strength of 10 or more	Certificate of good performance of machine/maintenance services for at least one year from TWO reputed government/private hospital Reason for Change: The performance certificate should not be bound to no of beds of nursery, because the user report is for unit.	No Change
5	Pg 1 Supplies Para 1	F & P MR 850 (including heater wire adaptor): 2 with each unit	Request to amend as integrated Humidifier of Same Make Justification: F&P is brand Specific Humidifier favouring single Company, for better and wider participation we request to amend as Integrated Humidifier of Same Make F&P MR 850 is name of specific company model. There are make in India is also available hence request to ma General request to change: Please delete specific name and model We request to amend as Inbuilt Humidifier of same make. F & P MR 850 or equivalent (including heater wire adaptor): 2 with each unit Reason for Change: The specific model of any company cannot be quoted for open bid. Hence it is requested to kindly allow the equivalent models.	No Change