

| Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022) | | | | |
|--|---|---|--|--------------------------|
| Bid Number: GEM/2022/B/2574936 Dated: 24-09-2022 | | | | |
| Item Name: Pulse Oximeter | | | | |
| Sr. No. | Buyer uploaded ATC document Page & Para | TENDER SPECIFICATION | REPRESENTATION RECEIVED FROM THE FIRM | COMMITTEE RECOMMENDATION |
| 1 | Pg 1 Para 6 | TFT screen- 6 inch wide. Visible from all sides. Option of displaying only HR and saturations. Should also display the AC status and Battery charge status | <p>The proposed solution should be 4 inch wide. Visible from all sides. Option of displaying only HR and saturations. Should also display the AC status and Battery charge status.</p> <p>Remark/Request/Suggestion: After amendment more and more company can participate in this tender.</p> <p>When the Government of India is promoting "AATMANIRBHAR BHARAT" & "MAKE IN INDIA" Public Procurement (preference to Make in India) order 2017</p> <p>We are looking your support to give chance to Indian Manufacturers'.</p> <p>TFT or LCD screen- 5 inch or wide. Visible from all sides. Option of displaying only HR and saturations. Should also display the AC status and Battery charge status</p> <p>Justification: For wider participation</p> | No Change |
| 2 | Pg 1 Para 16 | Should prevent ingress of water: IPX2 | <p>Should prevent ingress of water: IPX2 or other</p> <p>Justification: Please allow equivalent terms for wider participation</p> | No Change |
| 3 | Pg 1 Para 17 | <p>Equipment should be European CE with four digit notified body number or US FDA or BIS approved for the quoted model and certificate to be submitted.</p> <p>OR</p> <p>Should meet IEC 60601-1, IEC 60601-1-2 & ISO 80601-2-61 standards and should submit valid test report for the quoted model</p> | <p>Should be amended and replaced as</p> <p>IS 13450 Part 1 : 2018, IS 13450 Part 2 : Sec 49 : 2018, IS / ISO 80601 Part 2 : Sec 61 : 2011, IS 13450 Part 1, Sec 2 : 2012</p> <p>Justification: As per Government of India Notification</p> <p>1. Guidelines for Implementing provisions of Public Procurement (Preference to Make In India) Order (PPO), 2017 related to procurement of Goods & services in Medical Devices</p> <p>2. No. X 11035/379/2015-DFQC (Pt) Govt of India Ministry of Health & Family Welfare Dept of Health & Family Welfare</p> <p>3. No. 31026/36/2016-MD dtd. 16.10.2018</p> <p>Attaching the notification</p> | No Change |