

| Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022) | | | | |
|--|--|--------------------------------------|--|--------------------------|
| Bid Number: GEM/2022/B/2573846 Dated: 24-09-2022 | | | | |
| Item Name: High Frequency Ventilator (Oscillator) (PICU) | | | | |
| Sr. No. | Buyer Specification Document Page & Para | TENDER SPECIFICATION | REPRESENTATION RECEIVED FROM THE FIRM | COMMITTEE RECOMMENDATION |
| 1 | Pg 1 Para 3 a | Mean airway pressure: upto 45 cm H2O | Mean airway pressure: at least 40-45cm H2O Advantages/Benefits to Institutes after Change in Technical Specification: As it will leads to a wider participation of the bidders. | No Change |