

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575240 Dated: 24-09-2022				
Item Name: Radiant Warmer with Resuscitation Trolley				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 6	Mattress cover: removable with zipper, waterproof, washable, resistant to cleaning with chlorine-based solutions, and is flame retardant	The Unit should be supplied with superior quality completely Thermally Sealed Mattress to provide maximum patient safety and comfort. Justification: To avoid infection.	No Change
2	Pg 1 Para 8	Overhead light: Halogen or LED spot, with dimming function	Overhead light: Halogen or LED spot Justification: For wider Participation.	<b>Amended as:</b> Overhead light: Halogen or LED spot
3	Pg 1 Para 14	Inbuilt pulse oximeter: Should read oxygen saturation range, heart rate and plethysmograph; SET technology	To be deleted Justification: For wider Participation.	No Change
4	Pg 1 Para 15	Inbuilt flow driven T-piece resuscitator:	Flow driven T-piece resuscitator: Justification: For wider Participation	No Change
5	Pg 1 Para 15 d.	FiO2 concentration: 21 % to 100 %; blending with an electronic blender	FiO2 concentration: 21 % to 100 %; blending Justification: For wider participation FiO2 concentration: 21 % to 100 %; blending with an electronic/mechanical blender Reason for Changes: The mechanical dial in blender can be fixed with NRU and does same work. Also the option is available for warmers in India	No Change
6	Pg 1 Para 15 e.	Display to show delivered pressure, high/low pressure limits, oxygen concentration set value	Please delete this. Reason for Changes: the addition of Display for FiO2, hi low pressure limit shall make it favourable for imported goods. MII products will have certain limitations for this option	No Change
7	NA	Suggested by Bidder	The Unit should be supplied with superior quality completely Thermally Sealed Mattress to provide maximum patient safety and comfort. Justification: To avoid infection.	No Change
8	NA	Suggested by Bidder	The equipment should US FDA and CE certified. The supporting document in this regard should be submitted along with the bid. Justification: This is a safety standard feature for Equipment as well as patient.	No Change

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575240 Dated: 24-09-2022				
Item Name: Radiant Warmer with Resuscitation Trolley				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
9	Pg 2 Para 4	Warranty and CMC must include free replacement of:	Please delete free replacement Reason for Changes: The reasons are as mentioned below	No Change
10	Pg 2 Para 4 a	LED control panel	Reason for Changes: Consumable	No Change
11	Pg 2 Para 4 b	Side panels	Reason for Changes: Acrylic, normally damaged due to wear & tear	No Change
12	Pg 2 Para 4 c	Hinges, bends	Reason for Changes: Same as above	No Change
13	Pg 2 Para 4 d	Corner assembly	Reason for Changes: Same as above	No Change
14	Pg 2 Para 4 e	Locks and other attachments	Reason for Changes: Same as above	No Change
15	Pg 2 Para 4 f	Touch keys	Reason for Changes: Rough use only can damage the panel	No Change