

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575442 Dated: 24-09-2022				
Item Name: Phototherapy Units				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 3	Single head, surface size approx: 0.50 x 0.75 m	Single head, surface size approx: 0.35 to 0.50 x 0.23 to 0.75 m Justification: For wider participation	Amended as: Single head, surface size to cover the body surface of a term neonate (shall be verified during demo)
2	Pg 1 Para 4	Head height adjustable, approx: 1.40 to 1.75 m	Head height adjustable, approx: 1.13 to 1.6 m or more Justification: For wider participation	Amended as: Head height adjustable, approx: 1.10 to 1.75 m; the height adjustment mechanism should be sturdy and heavy duty (to be checked during demo)
3	Pg 1 Para 11	Should be BIS or European CE or US FDA approved product OR Should meet IEC 60601-1 , IEC 60601-1-2 and IEC 60601-2-50 standards and should submit valid test report from any NABL accredited lab or from the labs in their country of origin (in case of foreign manufacturers) for the quoted model	<p>The equipment should US FDA and CE certified. The supporting document in this regard should be submitted along with the bid. Justification: This is a safety standard feature for Equipment as well as patient.</p> <p>It should be amend with ISO 13485/BIS or CE Compliance/European CE or USFDA approved product Note: We have huge installation based at Govt sectors, Military hospital's as well as private hospitals. Recently we got order from HLL, Kerala with 56 nos. of Phototherapy and supplied and installed. FYR, attached PO copy</p> <p>Remarks/ Request/Suggestion: After amendment more and more company can participate in this tender. When the Government of India is promoting "AATMANIRBHAR BHARAT" & "MAKE IN INDIA" Public Procurement (preference to Make In India) order 2017 We are looking your support to give chance to Indian Manufacturers.</p>	No Change
4	NA	NA	Kindly confirm phototherapy machine Gem bid is for mobile stand or Double surface with bassinet	For mobile stand
5	NA	Suggested by Bidder	LED Phototherapy on mobile stand with height adjustment facility	Added Para: Should have (a) swivelling overhead unit and (b) robust, heavy duty joint (to be checked during demo)
6	NA	Suggested by Bidder	Should have swivelling overhead unit	

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575442 Dated: 24-09-2022				
Item Name: Phototherapy Units				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
7	NA	Suggested by Bidder	Should have ERTL certified Blue LEDs with wavelength of 465 – 510 nm and > 20,000 Hours Life	No Change
8	NA	Suggested by Bidder	Should have Irradiance 32 – 35 micro watt/cm2/nm	
9	NA	Suggested by Bidder	No UV and Infrared radiation should be there	
10	NA	Suggested by Bidder	No heat should be produced and thus dehydration is reduced	
11	NA	Suggested by Bidder	Should work on low power consumption 14W	No Change
12	NA	Suggested by Bidder	Should have negligible maintenance cost	
13	NA	Suggested by Bidder	Should be shock proof	
14	NA	Suggested by Bidder	Power Source : 230 V / 50 Hz	
15	NA	Suggested by Bidder	Should have irradiance : > 50uW/cm2/nm at a distance of 30 cm	No Change
16	NA	Suggested by Bidder	Should cover approximate area : 60cm x 30 cm	
17	NA	Suggested by Bidder	Should have digital hour meter to record the working : Total Usage of phototherapy	
18	NA	Suggested by Bidder	Should have 4 castors	