

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575529 Dated: 24-09-2022				
Item Name: High Frequency Ventilator (NICU)				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 3 a	Mean airway pressure: up to 45 cm H2O	Mean airway pressure: at least 40-45cm H2O Advantages/Benefits to Institutes after Change in Technical Specification: As it will leads to a wider participation of the bidders.	No Change