

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575138 Dated: 24-09-2022				
Item Name: Radiant Warmer (Stand-alone with Bassinet)				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 3	Heating element: quartz with parabolic reflector and protected by a metal grid; mattress temperature should not vary more than 10C from the centre to periphery (to be tested during the demo)	<p>Heating element: Calrod parabolic reflector protected by a metal grid and The Heating element should have lifetime warranty. Mattress temperature should not vary more than 10C from the center to periphery (to be tested during the demo) Justification: Calrod Heater creates 'FAR INFRARED' and not NEAR. (FAR INFRARED is absorbed by Cornea to prevent damage to Retina).</p> <p>Heating element: quartz /ceramic/calrod with parabolic reflector and protected by a metal grid; mattress temperature should not vary more than 10C from the centre to periphery (to be tested during the demo) Reason for Change: The Heating quality of Ceramic is better than quartz. The good quality encapsulated imported ceramic heaters are now available in India. which have longer life & are shatter proof & hence baby safe. Hence it is requested to kindly accept ceramic heaters also</p>	Amended as: Heating element: quartz or calrod with parabolic reflector and protected by a metal grid; mattress temperature should not vary more than 1°C (One degree Celcius) from the centre to periphery (to be tested during the demo)
2	Pg 1 Para 5	Overhead light: LED lamp that illuminates the baby area for procedures; adjustable intensity of light ranging up to 1000 lux or more	Overhead light: LED lamp that illuminates the baby area for procedures; intensity of light ranging up to 1000 lux or more Justification: For wider participation	No Change
3	Pg 1 Para 7	Heater output should switch off when swivelled from the midline; hood returns to midline with click	<p>Please change to "The heater swivel should be in such a way that it remains focussed on baby even in swivel condition"</p> <p>Reason for Change: The asked specs is contrary requirement as baby heating should not be discontinued even in procedures.</p>	No Change
4	Pg 1 Para 22	Manufacturer should have ISO 13485	<p>The equipment should US FDA and CE certified. The supporting document in this regard should be submitted along with the bid.</p> <p>Justification: This is a safety standard feature for Equipment as well as patient.</p>	No Change

Response To Pre-Bid Queries (Pre-Bid date: 30.09.2022)				
Bid Number: GEM/2022/B/2575138 Dated: 24-09-2022				
Item Name: Radiant Warmer (Stand-alone with Bassinet)				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
5	NA	Suggested by Bidder	The Unit should be supplied with superior quality completely Thermally Sealed Mattress to provide maximum patient safety and comfort. Justification: To avoid infection.	No Change
6	Pg 1 Supplies Para 2	Warranty and CMC must include free replacement of:	Please delete free replacement Reason for Change: The spares required as free replacement is not possible due to following reasons	No Change
7	Pg 1 Supplies Para 2 a	LED control panel	Reason for Change: Its consumable	No Change
8	Pg 2 Supplies Para 2 b	Side panels	Reason for Change: Its acrylic and normally break due to wear & tear	No Change
9	Pg 2 Supplies Para 2 c	Hinges, bends	Reason for Change: Same as above	No Change
10	Pg 2 Supplies Para 2 d	Corner assembly	Reason for Change: Same as above	No Change
11	Pg 2 Supplies Para 2 e	Locks and other attachments	Reason for Change: Same as above	No Change
12	Pg 2 Supplies Para 2 f	Touch keys	Reason for Change: Wrongly push can make it defective	No Change