

Amendment No. 1**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry No.: HITES/PCD/AIIMS-IV/63/Radiology/22-23 dated 03-03-2023**

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION – VII**TECHNICAL SPECIFICATIONS**

Item: CT 256 Slice (2023_HLL_146517_1)		
(Clause No & Page No)	Tender Specification	Technical Specification Committee (TSC) Recommendation
Point No 6 (f), Pg No. 46	Multi Slice CT Fluoroscopy to be quoted as standard. Price should be quoted separately	Multi Slice CT Fluoroscopy to be quoted as standard.
Point No 8 (g), Pg No. 48	Should have model based iterative reconstruction technique (ASiR-V /VEO/ADMIRE/IMR/ AIDR) for X Ray dose reduction. If VEO is not available, vendor should submit an undertaking from OEM. Added Para : Latest Deep Learning Algorithm for image reconstruction to be provided , if available with the vendor.	Should have model based iterative reconstruction technique (ASiR-V /VEO/ADMIRE/IMR/ AIDR) for X Ray dose reduction. If VEO is not available, vendor should submit an undertaking from OEM.

All other contents of the Tender enquiry including terms & conditions remain unaltered.

Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**
- II. Technical Compliance Sheet ‘Ver 1.0’ and Price bid Sheet ‘Ver 1.0’ in excel form against referred tender are being uploaded under ‘Corrigendum’ on CPPP. The same need to be downloaded for submission.**