

30-Apr-2022

**Amendment No. 02****Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry No. HITES/PCD/PMSSY-IV/19/Mix/2021-22 dated 16-03-2022 and Its Amendment No. 01 dated 18-Apr-2022.**

Following changes are being incorporated in the above referred Tender Enquiry Document.

**SECTION I**  
**NOTICE INVITING TENDER (NIT)**

Sl. No.	Tender ID	Short Description of goods	Department	
			For	Read As
6	2022_HLL_1 10085_6	Automated Microbial Identification and Sensitivity System	Neurosurgery	Microbiology

**For:**

(2) Tender Timeline:

Sl. No.	Description	Schedule
1	Last date and time of online submission of tender	05-05-2022 at 01:00 PM
2	Last date and time of physical submission of Tender Processing Fee, EMD or any other document specified in the Bidding Document	06-05-2022 at 02:00 PM
3	Date and time of tender Opening	06-05-2022 at 02:30 PM

**Read As:**

Sl. No.	Description	Schedule
1	Last date and time of online submission of tender	19-05-2022 at 01:00 PM
2	Last date and time of physical submission of Tender Processing Fee, EMD or any other document specified in the Bidding Document	20-05-2022 at 02:00 PM
3	Date and time of tender Opening	20-05-2022 at 02:30 PM

**SECTION - IV**  
**GENERAL CONDITIONS OF CONTRACT (GCC)**

Clause No.	For	Read As
21.1 A a) (i)	Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount	Four copies of supplier's invoice indicating Bill to Consignee through M/s HLL Infra Tech Services Limited and Ship to Place of supply,

		showing contract number, goods description, quantity, unit price and total amount. Invoices should be raised by the supplier in the name of consignee with their GSTN only
8. Inspection, Testing and Quality Control		<b>Added:</b> 8.9 Pre Despatch Inspection of goods located within India will be carried out by HITES/its designated representative or reputed third party agencies like SGS/ Lloyds/ TUV/ BV/ UL at manufacturers/ supplier's premises, and despatch clearance shall be issued by the respective Inspection Authority.

**SECTION - VII**  
**TECHNICAL SPECIFICATION**

Tender Para	TENDER SPECIFICATION	READ AS
<b>Item No. 01 Intra-Aortic Balloon Pump (IABP)</b>		
		<b>Added Para:</b> Bidder should quote the unit price of all the reusable & disposable accessories separately and the price should be frozen during warranty period. GST & all other taxes will be applicable at prevailing rates at the time of purchase.
		<b>Added para:</b> The unit shall be capable of being stored continuously in ambient temperature of 0 -50° C and relative humidity of 15-90%
		<b>Added para:</b> The unit shall be capable of operating in ambient temperature of 20-30° C and relative humidity of less than 70%.
<b>Item No. 02 - ACT Machine</b>		
Tender Para	TENDER SPECIFICATION	READ AS
3.1	ACT machine having at least two test well	ACT machine having at least <b>one test well</b>
3.2	The equipment should have 2 channel for measuring the individual clotting time and average difference for the two channels with the help of 2 channel test cartridge	<b>Deleted</b>
3.6	The display should show the coagulation time <b>of 2 channels</b>	The display should show the coagulation time
BOQ.2	ACT Tubes - 200 nos	ACT Tubes - <b>600 Nos</b>

Item No 03 Ultrasonic Surgical Aspirator/ Dissector (CUSA) (CTVS)		
Tender Para	TENDER SPECIFICATION	READ AS
	Other Accessories	Other Accessories
2	Suction Canister 1.5 to 2 L, Reusable, Qty. 10 no's.	Suction Canister 1.5 to 2 L, Reusable, Qty. 10 no's <b>OR</b> Disposable suction canister 1.5 to 2 L - qty 50 nos
BOQ.8	Suction Canister, Reusable - 10 Nos.	Suction Canister, Reusable - 10 Nos <b>OR</b> Disposable suction canister - 50 Nos
		<b>Added Para:</b> Bidder should quote the unit price of all the reusable & disposable accessories separately and the price should be frozen during warranty period. GST & all other taxes will be applicable at prevailing rates at the time of purchase.
Item No 04 Ultrasonic Surgical Aspirator (Neurosurgery)		
4.1.5	Cleaning brush for instrument lumen-2	<b>Cleaning brush/ Cleaning wire</b> for instrument lumen - 2 Nos.
BOQ.6	Cleaning brush for instrument lumen - 2 Nos	<b>Cleaning brush/ Cleaning wire</b> for instrument lumen - 2 Nos.
Item No 07 Flow Cytometer		
2	System should be equipped with solid state blue, Red & Violet lasers with at least 7 colours	2. System should be equipped with at least <b>3 solid state lasers (blue, Red &amp; Violet lasers)</b> with at least <b>10 colours</b> and <b>upgradability to 12 colours or more</b>
		<b>Added Para:</b> 5. System must have sample carryover of <0.1% or better
5	Compensation setting before, during and after data collection	6. Compensation setting before, during and after data collection
6	System should be quoted with both sampler options to fit in at least 20 tubes and 48/96 well microtiter plates, simultaneously or one at a time	7. System should be quoted with both sampler acquisition <b>options of single tube and at least 20 tubes carousel for auto acquisition. Bidder may offer optional 48 /96/384 well microtiter plates.</b>
7	System should be supported by appropriate software for acquisition & analysis of data	8. System should be supported by appropriate software for acquisition & analysis of data
8	System should have high resolution with 5-7 decade	9. System should have high resolution with 5-7 decade
		<b>Added Para:</b> 10. System must have LIS compatibility.
9	Touch/LCD screen computer with at least 21" monitor with colour printer to	11. Touch/LCD screen computer with at least <b>21" dual monitor</b> with at least 4core processor, 1TB

**HLL Infra Tech Services Limited**

	run and analyze the data (4core processor, 1TB HDD, 16 GB RAM or Higher).	HDD, 16 GB RAM or Higher. Bidder should provide at least one color laser printer. <b>Also should be installed with latest compatible windows OS and MS office.</b>
		<b>Added para:</b> 12. Bidder should provide at least one additional offline workstation with similar specifications and software as mentioned in point no 8 & 11 above
10	Compatible UPS with at least 1 hour back up should be supplied	13. Compatible UPS with at least 1 hour back up should be supplied.
11	Should be FDA or European CE or BIS approved product	14. Should be US FDA or <b>European CE- IVD</b> or BIS approved product
12	List of users and track record of installation of similar equipment should be provided.	15. List of users and track record of <b>at least 5 installation of quoted model in past 5 years</b> should be provided.
		<b>Added para:</b> 16. The necessary reagents for functioning of flowcytometer like sheath fluids, shutdown fluids, QC beads etc. (10 nos. each). These are to be supplied at the time of installation except QC beads which will be as per the requirement of user.  Bidder should quote the unit price separately and price should be valid for warranty period.
		<b>Added para:</b> 17. The bidder should give commitment for onsite training of personnel and regular technical support as and when required.

**Note:**

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.
- ii. All other contents of the tender enquiry including terms & conditions remain unaltered.
- iii. TC compliance sheet v1.0 and Price bid format v1.0 is being uploaded on CPP Portal for necessary actions of interested Bidders.