

Response To Pre-Bid Queries			
GEM/2022/B/2117164			
Harmonic Scalpel			
Tender Para	TENDER SPECIFICATION	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
	Harmonic Scalpel	We request you to kindly amend it as Ultrasonic & Advance Bipolar device Reason: Harmonic Scalpel is Trade name of one specific company	Ameded as: Ultrasoic cutting annd coagulation system with advaced bipolar device
	Accessories:		
4	Ultrasonic 5 mm Laparoscopic hand activated curved tip coagulating probe Quantity 2 nos	We request you to kindly amend it as integrated RF/bipolar or ultrasonic energy hand activated curved tip coagulating probe for Lap surgery Reason: Olympus has Integrated Ultrasonic & RF Energy in Single probe	No change considered
5	Ultrasonic hand activated curved tip coagulating probe for open surgery: Quantity 2nos	We request you to kindly amend it as integrated RF/bipolar or ultrasonic energy hand activated curved tip coagulating probe for Open surgery Reason: Olympus has Integrated Ultrasonic & RF Energy in Single probe	No change considered
	BOQ		
3	Hand piece/Hand pieces (transducer) compatible with various hand instruments used in open and laparoscopic surgeries. Sufficient number of transducers so as to last 90 cases to be supplied	We request you to kindly amend it as Hand piece/Hand pieces (transducer) compatible with various hand instruments used in open and laparoscopic surgeries should be reusable and autoclavable for min. 90 autoclave cycles. Remarks: We suggest to ask bidders to submit document from OEM for certifying autoclave life of Hand Piece(Transducer)	No change considered