

Response To Pre-Bid Queries			
GEM/2022/B/2092463 Dated- 07-04-2022			
ICU Ventilator			
Tender Para	TENDER SPECIFICATION	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
2	Screen should be minimum of 12" inch or more and integrated, tiltable single screen	Request to amend Screen Size minimum 15" color touchscreen <b>Reason:</b> We need to have monitoring of 3 waveforms and 2 loops simultaneously which can be seen and an analyzed on a bigger screen	No change considered
5.a	Should have the facility for following settings: Tidal Volume: Minimum 5ml or less and maximum of 1500 ml or more in Volume control	Should have the facility for following settings: Tidal Volume: Minimum 20 ml or less and maximum of 1500 ml or more in Volume control	No change considered
		Should have the facility for following settings: Tidal Volume: Minimum 20 ml or less and maximum of 1500 ml or more in Volume control	No change considered
13(b)	Ventilator should have external compressor, from the same manufacturer (Price to be quoted separately)	Request to amend Compressor as OEM Integrated <b>Reason:</b> To have better uniformity of product and timely service backup from single company we request to amend Compressor as OEM integrated.	No change considered